

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034902.D
 Acq On : 01 Apr 2023 05:00
 Operator : JC/MD
 Sample : 02137-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-3D

Quant Time: Apr 03 03:20:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	272336	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	468064	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	417799	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181447	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	214117	51.409	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.820%
35) Dibromofluoromethane	5.385	113	147935	47.785	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.560%
50) Toluene-d8	8.647	98	566486	49.574	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.140%
62) 4-Bromofluorobenzene	11.079	95	212241	47.997	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034902.D
Acq On : 01 Apr 2023 05:00
Operator : JC/MD
Sample : 02137-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-3D

Quant Time: Apr 03 03:20:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

