

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034903.D
 Acq On : 01 Apr 2023 05:23
 Operator : JC/MD
 Sample : 02137-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-2R

Quant Time: Apr 03 03:21:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	262608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	450682	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	391873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	161081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	204563	50.934	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.860%
35) Dibromofluoromethane	5.385	113	143384	48.101	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.200%
50) Toluene-d8	8.647	98	537379	48.841	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.079	95	192764	45.274	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.540%
Target Compounds						
						Qvalue
3) Chloromethane	1.295	50	749	0.228	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034903.D
Acq On : 01 Apr 2023 05:23
Operator : JC/MD
Sample : 02137-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-2R

Quant Time: Apr 03 03:21:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

