

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034906.D
 Acq On : 01 Apr 2023 06:33
 Operator : JC/MD
 Sample : 02137-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK-1

Quant Time: Apr 03 02:10:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	266891	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	456884	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	410070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	180822	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	207177	50.757	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.520%	
35) Dibromofluoromethane	5.385	113	143904	47.620	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.240%	
50) Toluene-d8	8.647	98	551877	49.478	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.960%	
62) 4-Bromofluorobenzene	11.079	95	209760	48.597	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.200%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034906.D
Acq On : 01 Apr 2023 06:33
Operator : JC/MD
Sample : 02137-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK-1

Quant Time: Apr 03 02:10:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

