

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034907.D
 Acq On : 01 Apr 2023 06:56
 Operator : JC/MD
 Sample : 02137-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Apr 03 02:10:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	272478	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	469367	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	415185	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	182997	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	210143	50.428	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.860%
35) Dibromofluoromethane	5.385	113	147522	47.519	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.040%
50) Toluene-d8	8.647	98	559568	48.833	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.660%
62) 4-Bromofluorobenzene	11.079	95	211801	47.764	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.520%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	9567	5.651	ug/l	89
20) Methylene Chloride	2.788	84	4943	1.480	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034907.D
Acq On : 01 Apr 2023 06:56
Operator : JC/MD
Sample : 02137-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Apr 03 02:10:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

