

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034912.D
 Acq On : 01 Apr 2023 08:52
 Operator : JC/MD
 Sample : 02159-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MP-5

Quant Time: Apr 03 02:12:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	277731	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	478318	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	423064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	182581	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	212530	50.037	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.080%
35) Dibromofluoromethane	5.385	113	151267	47.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.620%
50) Toluene-d8	8.647	98	573478	49.110	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.220%
62) 4-Bromofluorobenzene	11.079	95	213699	47.291	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.580%
Target Compounds						
						Qvalue
64) Tetrachloroethene	9.275	164	1004	0.332	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034912.D
Acq On : 01 Apr 2023 08:52
Operator : JC/MD
Sample : 02159-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MP-5

Quant Time: Apr 03 02:12:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

