

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034915.D
 Acq On : 01 Apr 2023 10:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 03 02:13:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.499	0.552	-10.6	110	0.00
3 P	Chloromethane	0.624	0.639	-2.4	99	0.00
4 C	Vinyl Chloride	0.618	0.623	-0.8#	99	0.00
5 T	Bromomethane	0.402	0.334	16.9	81	0.00
6 T	Chloroethane	0.402	0.355	11.7	88	0.00
7 T	Trichlorofluoromethane	0.994	0.885	11.0	87	0.00
8 T	Diethyl Ether	0.350	0.345	1.4	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.519	0.509	1.9	95	0.00
10 T	Methyl Iodide	0.647	0.570	11.9	81	0.00
11 T	Tert butyl alcohol	0.134	0.123	8.2	85	-0.01
12 CM	1,1-Dichloroethene	0.513	0.487	5.1#	92	0.00
13 T	Acrolein	0.138	0.118	14.5	82	0.00
14 T	Allyl chloride	1.006	0.971	3.5	93	0.00
15 T	Acrylonitrile	0.302	0.319	-5.6	100	0.00
16 T	Acetone	0.311	0.300	3.5	93	0.00
17 T	Carbon Disulfide	1.286	1.148	10.7	84	0.00
18 T	Methyl Acetate	1.104	1.189	-7.7	104	0.00
19 T	Methyl tert-butyl Ether	1.900	1.902	-0.1	96	0.00
20 T	Methylene Chloride	0.613	0.578	5.7	95	0.00
21 T	trans-1,2-Dichloroethene	0.560	0.539	3.8	93	0.00
22 T	Diisopropyl ether	2.029	2.045	-0.8	97	0.00
23 T	Vinyl Acetate	1.447	1.509	-4.3	95	0.00
24 P	1,1-Dichloroethane	1.092	1.086	0.5	96	0.00
25 T	2-Butanone	0.452	0.464	-2.7	95	0.00
26 T	2,2-Dichloropropane	0.898	0.632	29.6#	66	0.00
27 T	cis-1,2-Dichloroethene	0.663	0.642	3.2	94	0.00
28 T	Bromochloromethane	0.534	0.516	3.4	92	0.00
29 T	Tetrahydrofuran	0.283	0.303	-7.1	100	0.00
30 C	Chloroform	1.109	1.108	0.1#	95	0.00
31 T	Cyclohexane	0.931	0.940	-1.0	94	0.00
32 T	1,1,1-Trichloroethane	0.946	0.957	-1.2	94	0.00
33 S	1,2-Dichloroethane-d4	0.765	0.785	-2.6	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.331	0.325	1.8	92	0.00
36 T	1,1-Dichloropropene	0.490	0.474	3.3	95	0.00
37 T	Ethyl Acetate	0.512	0.534	-4.3	97	0.00
38 T	Carbon Tetrachloride	0.462	0.472	-2.2	96	0.00
39 T	Methylcyclohexane	0.520	0.488	6.2	91	0.00
40 TM	Benzene	1.406	1.347	4.2	94	0.00
41 T	Methacrylonitrile	0.279	0.305	-9.3	102	0.00
42 TM	1,2-Dichloroethane	0.559	0.564	-0.9	98	0.00
43 T	Isopropyl Acetate	0.857	0.865	-0.9	96	0.00
44 TM	Trichloroethene	0.359	0.343	4.5	93	0.00
45 C	1,2-Dichloropropane	0.367	0.355	3.3#	94	0.00
46 T	Dibromomethane	0.252	0.243	3.6	93	0.00
47 T	Bromodichloromethane	0.466	0.476	-2.1	95	0.00
48 T	Methyl methacrylate	0.421	0.445	-5.7	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034915.D
 Acq On : 01 Apr 2023 10:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 03 02:13:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	89	-0.01
50 S	Toluene-d8	1.221	1.197	2.0	91	0.00
51 T	4-Methyl-2-Pentanone	0.507	0.534	-5.3	99	0.00
52 CM	Toluene	0.917	0.855	6.8#	94	0.00
53 T	t-1,3-Dichloropropene	0.501	0.478	4.6	84	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.526	4.0	86	0.00
55 T	1,1,2-Trichloroethane	0.347	0.340	2.0	94	0.00
56 T	Ethyl methacrylate	0.544	0.565	-3.9	94	0.00
57 T	1,3-Dichloropropane	0.615	0.605	1.6	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.277	0.294	-6.1	98	0.00
59 T	2-Hexanone	0.384	0.404	-5.2	97	0.00
60 T	Dibromochloromethane	0.331	0.338	-2.1	92	0.00
61 T	1,2-Dibromoethane	0.361	0.359	0.6	94	0.00
62 S	4-Bromofluorobenzene	0.472	0.483	-2.3	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.358	0.322	10.1	90	0.00
65 PM	Chlorobenzene	1.040	0.987	5.1	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.358	0.359	-0.3	92	0.00
67 C	Ethyl Benzene	1.885	1.832	2.8#	94	0.00
68 T	m/p-Xylenes	0.720	0.686	4.7	93	0.00
69 T	o-Xylene	0.705	0.682	3.3	94	0.00
70 T	Styrene	1.126	1.134	-0.7	94	0.00
71 P	Bromoform	0.243	0.247	-1.6	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.663	3.601	1.7	94	0.00
74 T	N-amyl acetate	1.647	1.750	-6.3	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.200	1.198	0.2	96	0.00
76 T	1,2,3-Trichloropropane	1.007	1.006	0.1	95	0.00
77 T	Bromobenzene	0.908	0.841	7.4	91	0.00
78 T	n-propylbenzene	4.231	4.127	2.5	93	0.00
79 T	2-Chlorotoluene	2.671	2.593	2.9	93	0.00
80 T	1,3,5-Trimethylbenzene	3.062	3.033	0.9	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.336	0.281	16.4	75	0.00
82 T	4-Chlorotoluene	3.102	3.003	3.2	93	0.00
83 T	tert-Butylbenzene	2.876	2.822	1.9	93	0.00
84 T	1,2,4-Trimethylbenzene	3.114	3.030	2.7	93	0.00
85 T	sec-Butylbenzene	3.498	3.412	2.5	93	0.00
86 T	p-Isopropyltoluene	2.917	2.881	1.2	92	0.00
87 T	1,3-Dichlorobenzene	1.653	1.602	3.1	92	0.00
88 T	1,4-Dichlorobenzene	1.726	1.621	6.1	91	0.00
89 T	n-Butylbenzene	2.549	2.432	4.6	89	0.00
90 T	Hexachloroethane	0.452	0.457	-1.1	88	0.00
91 T	1,2-Dichlorobenzene	1.655	1.599	3.4	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.251	0.274	-9.2	94	0.00
93 T	1,2,4-Trichlorobenzene	0.936	0.917	2.0	91	0.00
94 T	Hexachlorobutadiene	0.412	0.343	16.7	86	0.00
95 T	Naphthalene	3.326	3.354	-0.8	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034915.D
Acq On : 01 Apr 2023 10:01
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 03 02:13:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.944	0.904	4.2	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6