



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034862.D
Acq On : 31 Mar 2023 08:23
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Title : SW846 8260
Last Update : Sat Apr 01 01:50:18 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.8	16504	PASS
75	95	30	60	54.6	41309	PASS
95	95	100	100	100.0	75675	PASS
96	95	5	9	6.7	5099	PASS
173	174	0.00	2	1.0	531	PASS
174	95	50	100	71.9	54379	PASS
175	174	5	9	8.0	4336	PASS
176	174	95	101	97.0	52765	PASS
177	176	5	9	6.9	3641	PASS