

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040119\
 Data File : VX008561.D
 Acq On : 01 Apr 2019 11:54
 Operator : JC/SP
 Sample : VX0401WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0401WBL01

Quant Time: Apr 02 01:27:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	206508	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	341874	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	289722	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	121017	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	142621	45.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.12%	
35) Dibromofluoromethane	5.49	113	105437	47.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.60%	
50) Toluene-d8	8.71	98	423794	47.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.08%	
62) 4-Bromofluorobenzene	11.13	95	141973	42.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.96%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040119\
Data File : VX008561.D
Acq On : 01 Apr 2019 11:54
Operator : JC/SP
Sample : VX0401WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0401WBL01

Quant Time: Apr 02 01:27:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

