

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040119\
 Data File : VX008563.D
 Acq On : 01 Apr 2019 13:04
 Operator : JC/SP
 Sample : K2189-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201R-TB-20190329

Quant Time: Apr 02 01:33:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	219052	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	365302	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	316097	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133006	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	153911	46.35	ug/l	0.00
Spiked Amount			Recovery	=	92.70%	
35) Dibromofluoromethane	5.49	113	112728	47.33	ug/l	0.00
Spiked Amount			Recovery	=	94.66%	
50) Toluene-d8	8.71	98	455078	47.78	ug/l	0.00
Spiked Amount			Recovery	=	95.56%	
62) 4-Bromofluorobenzene	11.13	95	155351	44.02	ug/l	0.00
Spiked Amount			Recovery	=	88.04%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040119\
Data File : VX008563.D
Acq On : 01 Apr 2019 13:04
Operator : JC/SP
Sample : K2189-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201R-TB-20190329

Quant Time: Apr 02 01:33:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

