

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040119\
 Data File : VX008564.D
 Acq On : 01 Apr 2019 13:27
 Operator : JC/SP
 Sample : K2189-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201R-EB-20190329

Quant Time: Apr 02 01:40:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	224600	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	376358	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	316724	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	124587	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	156074	45.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.68%	
35) Dibromofluoromethane	5.50	113	115098	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
50) Toluene-d8	8.71	98	467798	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
62) 4-Bromofluorobenzene	11.13	95	152496	41.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.88%	
Target Compounds						
16) Acetone	2.44	43	10722	6.025	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040119\
Data File : VX008564.D
Acq On : 01 Apr 2019 13:27
Operator : JC/SP
Sample : K2189-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201R-EB-20190329

Quant Time: Apr 02 01:40:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

