

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040119\
 Data File : VX008566.D
 Acq On : 01 Apr 2019 14:14
 Operator : JC/SP
 Sample : K2189-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201R-GW-778-780

Quant Time: Apr 02 01:46:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	220200	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	368480	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	315797	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	121967	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	154097	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
35) Dibromofluoromethane	5.49	113	113169	47.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.20%	
50) Toluene-d8	8.71	98	461607	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
62) 4-Bromofluorobenzene	11.13	95	152284	42.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.56%	
Target Compounds						
16) Acetone	2.44	43	24248	13.898	ug/l	99
25) 2-Butanone	4.68	43	9449	3.378	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040119\
Data File : VX008566.D
Acq On : 01 Apr 2019 14:14
Operator : JC/SP
Sample : K2189-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201R-GW-778-780

Quant Time: Apr 02 01:46:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

