

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040119\
 Data File : VX008567.D
 Acq On : 01 Apr 2019 14:37
 Operator : JC/SP
 Sample : K2189-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201R-GW-798-800

Quant Time: Apr 02 01:48:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	233921	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	390114	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	336960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	141619	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	162635	45.86	ug/l	0.00
Spiked Amount			Recovery	=	91.72%	
35) Dibromofluoromethane	5.50	113	119201	46.86	ug/l	0.00
Spiked Amount			Recovery	=	93.72%	
50) Toluene-d8	8.71	98	490230	48.20	ug/l	0.00
Spiked Amount			Recovery	=	96.40%	
62) 4-Bromofluorobenzene	11.13	95	166158	44.08	ug/l	0.00
Spiked Amount			Recovery	=	88.16%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040119\
Data File : VX008567.D
Acq On : 01 Apr 2019 14:37
Operator : JC/SP
Sample : K2189-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201R-GW-798-800

Quant Time: Apr 02 01:48:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

