

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040119\
 Data File : VX008577.D
 Acq On : 01 Apr 2019 18:30
 Operator : JC/SP
 Sample : K2210-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WE-3

Quant Time: Apr 02 02:16:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	219706	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	363731	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	309042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	122365	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	153776	46.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.34%	
35) Dibromofluoromethane	5.50	113	113490	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
50) Toluene-d8	8.71	98	456819	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
62) 4-Bromofluorobenzene	11.13	95	149802	42.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.26%	
Target Compounds						
16) Acetone	2.44	43	8286	4.760	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040119\
Data File : VX008577.D
Acq On : 01 Apr 2019 18:30
Operator : JC/SP
Sample : K2210-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WE-3

Quant Time: Apr 02 02:16:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

