


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040119\  
Data File : VX008559.D  
Acq On : 01 Apr 2019 09:16  
Operator : JC/SP  
Sample : BFB  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X032519W.M  
Title : SW846 8260  
Last Update : Tue Mar 26 07:57:20 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 20.8      | 12638   | PASS             |
| 75          | 95           | 30           | 60           | 50.8      | 30870   | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 60725   | PASS             |
| 96          | 95           | 5            | 9            | 6.6       | 3999    | PASS             |
| 173         | 174          | 0.00         | 2            | 1.2       | 560     | PASS             |
| 174         | 95           | 50           | 100          | 74.2      | 45032   | PASS             |
| 175         | 174          | 5            | 9            | 7.5       | 3377    | PASS             |
| 176         | 174          | 95           | 101          | 96.9      | 43642   | PASS             |
| 177         | 176          | 5            | 9            | 6.4       | 2814    | PASS             |