


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX040120\  
 Data File : VX015457.D  
 Acq On : 01 Apr 2020 09:54  
 Operator : JC/SP  
 Sample : BFB  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X032720W.M  
 Title : SW846 8260  
 Last Update : Fri Mar 27 09:35:51 2020



AutoFind: Scans 1678, 1679, 1680; Background Corrected with Scan 1671

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 20.4      | 53548   | PASS             |
| 75          | 95           | 30           | 60           | 49.6      | 130369  | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 262762  | PASS             |
| 96          | 95           | 5            | 9            | 6.5       | 16998   | PASS             |
| 173         | 174          | 0.00         | 2            | 0.5       | 1094    | PASS             |
| 174         | 95           | 50           | 100          | 78.5      | 206187  | PASS             |
| 175         | 174          | 5            | 9            | 7.9       | 16389   | PASS             |
| 176         | 174          | 95           | 101          | 100.6     | 207523  | PASS             |
| 177         | 176          | 5            | 9            | 6.2       | 12815   | PASS             |