

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021708.D
 Acq On : 01 Apr 2021 15:17
 Operator : JC/MD
 Sample : PB135474TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135474TB

Quant Time: Apr 02 03:28:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	56666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	102403	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	114550	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	51977	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	49030	52.00	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.00%	
35) Dibromofluoromethane	5.470	113	38102	52.38	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.76%	
50) Toluene-d8	8.694	98	138708	53.51	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.02%	
62) 4-Bromofluorobenzene	11.118	95	55431	55.00	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.00%	
Target Compounds						
						Qvalue
16) Acetone	2.423	43	9933	27.86	ug/l	89
20) Methylene Chloride	2.835	84	3107	4.47	ug/l #	78
25) 2-Butanone	4.641	43	3282	6.07	ug/l #	83
43) Isopropyl Acetate	6.417	43	9232	4.90	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
Data File : VX021708.D
Acq On : 01 Apr 2021 15:17
Operator : JC/MD
Sample : PB135474TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135474TB

Quant Time: Apr 02 03:28:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

