

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021713.D
 Acq On : 01 Apr 2021 17:03
 Operator : JC/MD
 Sample : PB135474ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135474ZHE#01

Quant Time: Apr 02 03:29:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	54536	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	98646	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	109489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48911	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	47356	52.19	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.38%		
35) Dibromofluoromethane	5.464	113	36920	52.69	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.38%		
50) Toluene-d8	8.694	98	133931	53.64	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 107.28%		
62) 4-Bromofluorobenzene	11.118	95	52513	54.09	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 108.18%		
Target Compounds						
16) Acetone	2.423	43	9544	27.82	ug/l	91
20) Methylene Chloride	2.840	84	3215	4.80	ug/l #	85
25) 2-Butanone	4.641	43	2896	5.57	ug/l #	82
43) Isopropyl Acetate	6.411	43	8783	4.84	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
Data File : VX021713.D
Acq On : 01 Apr 2021 17:03
Operator : JC/MD
Sample : PB135474ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135474ZHE#01

Quant Time: Apr 02 03:29:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

