

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021714.D
 Acq On : 01 Apr 2021 17:27
 Operator : JC/MD
 Sample : PB135474ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135474ZHE#02

Quant Time: Apr 02 03:30:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	57019	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	106410	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	116737	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	51500	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51253	54.02	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.04%
35) Dibromofluoromethane	5.464	113	40064	53.00	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.00%
50) Toluene-d8	8.694	98	141043	52.36	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.72%
62) 4-Bromofluorobenzene	11.118	95	57182	54.60	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.20%
Target Compounds						
16) Acetone	2.423	43	10087	28.12	ug/l	Qvalue 89
20) Methylene Chloride	2.829	84	3214	4.59	ug/l #	82
25) 2-Butanone	4.641	43	2951	5.43	ug/l	94
43) Isopropyl Acetate	6.412	43	8900	4.55	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
Data File : VX021714.D
Acq On : 01 Apr 2021 17:27
Operator : JC/MD
Sample : PB135474ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135474ZHE#02

Quant Time: Apr 02 03:30:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

