

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021716.D
 Acq On : 01 Apr 2021 18:14
 Operator : JC/MD
 Sample : M1895-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Apr 02 03:30:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	58385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	107121	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	112459	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	51456	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	49593	51.05	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.10%	
35) Dibromofluoromethane	5.464	113	38807	51.00	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.00%	
50) Toluene-d8	8.694	98	135090	49.82	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.64%	
62) 4-Bromofluorobenzene	11.118	95	54806	51.99	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.98%	
Target Compounds						Qvalue
11) Tert butyl alcohol	3.017	59	196190	1221.61	ug/l	97
16) Acetone	2.423	43	2064	5.62	ug/l	94
19) Methyl tert-butyl Ether	3.164	73	1123772	511.44	ug/l	98
22) Diisopropyl ether	3.823	45	16908	7.29	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
Data File : VX021716.D
Acq On : 01 Apr 2021 18:14
Operator : JC/MD
Sample : M1895-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Apr 02 03:30:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

