

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021722.D
 Acq On : 01 Apr 2021 20:34
 Operator : JC/MD
 Sample : M1864-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MMR-SP

Quant Time: Apr 02 03:33:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	60960	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	110993	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	117808	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	51425	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50698	49.98	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.96%
35) Dibromofluoromethane	5.464	113	38718	49.11	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.22%
50) Toluene-d8	8.694	98	143755	51.17	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.34%
62) 4-Bromofluorobenzene	11.118	95	57179	52.35	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.70%
Target Compounds						
					Qvalue	
16) Acetone	2.423	43	9080	23.68	ug/l #	86
20) Methylene Chloride	2.834	84	3366	4.50	ug/l #	83
25) 2-Butanone	4.641	43	3143	5.41	ug/l #	84
43) Isopropyl Acetate	6.411	43	7698	3.77	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
Data File : VX021722.D
Acq On : 01 Apr 2021 20:34
Operator : JC/MD
Sample : M1864-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MMR-SP

Quant Time: Apr 02 03:33:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

