

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021723.D
 Acq On : 01 Apr 2021 20:58
 Operator : JC/MD
 Sample : M1776-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 110

Quant Time: Apr 02 03:33:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	59529	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	108070	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	121544	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	59163	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50141	50.62	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.24%			
35) Dibromofluoromethane	5.464	113	38411	50.04	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.08%			
50) Toluene-d8	8.694	98	140462	51.35	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.70%			
62) 4-Bromofluorobenzene	11.118	95	63498	59.70	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 119.40%			
Target Compounds					Qvalue	
11) Tert butyl alcohol	3.005	59	1593	9.73	ug/l #	91
16) Acetone	2.417	43	1001775	2674.94	ug/l	89
18) Methyl Acetate	2.746	43	1516	1.40	ug/l #	81
20) Methylene Chloride	2.834	84	5899	8.07	ug/l #	83
25) 2-Butanone	4.634	43	8293	14.61	ug/l	90
43) Isopropyl Acetate	6.411	43	9123	4.59	ug/l #	90
52) Toluene	8.764	92	79735	42.93	ug/l	98
64) Tetrachloroethene	9.317	164	4523	5.56	ug/l #	78
67) Ethyl Benzene	10.235	91	288564	64.84	ug/l	100
68) m/p-Xylenes	10.341	106	598505	377.47	ug/l	93
69) o-Xylene	10.682	106	306971	199.88	ug/l	93
73) Isopropylbenzene	11.000	105	7516	1.81	ug/l	99
78) n-propylbenzene	11.341	91	19035	3.88	ug/l	98
80) 1,3,5-Trimethylbenzene	11.488	105	35023	9.86	ug/l	99
84) 1,2,4-Trimethylbenzene	11.788	105	176709	49.36	ug/l	98
85) sec-Butylbenzene	11.930	105	6636	1.66	ug/l	87
86) p-Isopropyltoluene	12.047	119	4673	1.29	ug/l	88
95) Naphthalene	13.818	128	95388	24.73	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
Data File : VX021723.D
Acq On : 01 Apr 2021 20:58
Operator : JC/MD
Sample : M1776-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
110

Quant Time: Apr 02 03:33:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

