

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021725.D
 Acq On : 01 Apr 2021 21:45
 Operator : JC/MD
 Sample : M1895-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: Apr 02 03:34:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	61847	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	112755	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	116693	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52713	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51279	49.83	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.66%
35) Dibromofluoromethane	5.464	113	39841	49.74	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.48%
50) Toluene-d8	8.694	98	142414	49.90	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.80%
62) 4-Bromofluorobenzene	11.118	95	59224	53.37	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.74%
Target Compounds						
					Qvalue	
16) Acetone	2.423	43	1733	4.45	ug/l #	79
19) Methyl tert-butyl Ether	3.170	73	7630	3.28	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
Data File : VX021725.D
Acq On : 01 Apr 2021 21:45
Operator : JC/MD
Sample : M1895-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: Apr 02 03:34:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

