

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021734.D
 Acq On : 02 Apr 2021 11:50
 Operator : JC/MD
 Sample : M1895-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1R

Quant Time: Apr 03 02:02:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	57188	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	103745	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	103238	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	43990	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	48838	51.33	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.66%	
35) Dibromofluoromethane	5.470	113	37329	50.65	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.30%	
50) Toluene-d8	8.694	98	130099	49.54	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.08%	
62) 4-Bromofluorobenzene	11.118	95	50224	49.19	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.38%	
Target Compounds						Qvalue
16) Acetone	2.417	43	1376	3.82	ug/l	93
19) Methyl tert-butyl Ether	3.170	73	6921	3.22	ug/l	99
40) Benzene	6.111	78	7533	2.61	ug/l	97
52) Toluene	8.765	92	1725	0.97	ug/l	95
69) o-Xylene	10.682	106	445	0.34	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
Data File : VX021734.D
Acq On : 02 Apr 2021 11:50
Operator : JC/MD
Sample : M1895-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1R

Quant Time: Apr 03 02:02:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

