

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021700.D
 Acq On : 01 Apr 2021 11:49
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 03:24:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.550	0.496	9.8	85	0.00
3 P	Chloromethane	0.702	0.550	21.7	86	0.00
4 C	Vinyl Chloride	0.562	0.523	6.9#	87	0.00
5 T	Bromomethane	0.226	0.241	-6.6	104	0.00
6 T	Chloroethane	0.357	0.338	5.3	90	0.00
7 T	Trichlorofluoromethane	0.914	0.886	3.1	90	0.00
8 T	Diethyl Ether	0.317	0.293	7.6	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.515	0.500	2.9	93	0.00
10 T	Methyl Iodide	0.362	0.425	-17.4	111	0.00
11 T	Tert butyl alcohol	0.138	0.113	18.1	81	0.00
12 CM	1,1-Dichloroethene	0.504	0.443	12.1#	87	0.00
13 T	Acrolein	0.077	0.060	22.1	86	0.00
14 T	Allyl chloride	1.023	0.909	11.1	85	0.00
15 T	Acrylonitrile	0.327	0.299	8.6	86	0.00
16 T	Acetone	0.315	0.257	18.4	80	0.00
17 T	Carbon Disulfide	1.415	1.070	24.4	75	0.00
18 T	Methyl Acetate	0.909	0.926	-1.9	94	0.00
19 T	Methyl tert-butyl Ether	1.882	1.771	5.9	88	0.00
20 T	Methylene Chloride	0.614	0.558	9.1	89	0.00
21 T	trans-1,2-Dichloroethene	0.553	0.505	8.7	87	0.00
22 T	Diisopropyl ether	1.987	1.903	4.2	90	0.00
23 T	Vinyl Acetate	1.681	1.533	8.8	84	0.00
24 P	1,1-Dichloroethane	1.071	1.014	5.3	90	0.00
25 T	2-Butanone	0.477	0.417	12.6	83	0.00
26 T	2,2-Dichloropropane	0.928	0.881	5.1	89	0.00
27 T	cis-1,2-Dichloroethene	0.643	0.592	7.9	87	0.00
28 T	Bromochloromethane	0.536	0.481	10.3	91	0.00
29 T	Tetrahydrofuran	0.309	0.279	9.7	86	0.00
30 C	Chloroform	1.131	1.106	2.2#	92	0.00
31 T	Cyclohexane	0.915	0.811	11.4	82	0.00
32 T	1,1,1-Trichloroethane	1.002	0.955	4.7	89	0.00
33 S	1,2-Dichloroethane-d4	0.832	0.738	11.3	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.355	0.353	0.6	93	0.00
36 T	1,1-Dichloropropene	0.490	0.493	-0.6	89	0.00
37 T	Ethyl Acetate	0.566	0.531	6.2	85	0.00
38 T	Carbon Tetrachloride	0.526	0.557	-5.9	91	0.00
39 T	Methylcyclohexane	0.543	0.540	0.6	84	0.00
40 TM	Benzene	1.391	1.423	-2.3	88	0.00
41 T	Methacrylonitrile	0.329	0.318	3.3	86	-0.01
42 TM	1,2-Dichloroethane	0.603	0.626	-3.8	90	0.00
43 T	Isopropyl Acetate	0.920	0.899	2.3	84	0.00
44 TM	Trichloroethene	0.367	0.379	-3.3	91	0.00
45 C	1,2-Dichloropropane	0.377	0.393	-4.2#	89	0.00
46 T	Dibromomethane	0.275	0.281	-2.2	88	0.00
47 T	Bromodichloromethane	0.559	0.595	-6.4	91	0.00
48 T	Methyl methacrylate	0.491	0.489	0.4	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021700.D
 Acq On : 01 Apr 2021 11:49
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 03:24:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	85	0.00
50 S	Toluene-d8	1.266	1.257	0.7	90	0.00
51 T	4-Methyl-2-Pentanone	0.566	0.589	-4.1	88	0.00
52 CM	Toluene	0.859	0.908	-5.7#	89	0.00
53 T	t-1,3-Dichloropropene	0.564	0.592	-5.0	88	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.636	-5.8	89	0.00
55 T	1,1,2-Trichloroethane	0.363	0.391	-7.7	92	0.00
56 T	Ethyl methacrylate	0.548	0.557	-1.6	83	0.00
57 T	1,3-Dichloropropane	0.621	0.656	-5.6	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.307	0.313	-2.0	86	0.00
59 T	2-Hexanone	0.425	0.444	-4.5	88	0.00
60 T	Dibromochloromethane	0.406	0.461	-13.5	95	0.00
61 T	1,2-Dibromoethane	0.390	0.407	-4.4	88	0.00
62 S	4-Bromofluorobenzene	0.492	0.495	-0.6	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.335	0.402	-20.0	104	0.00
65 PM	Chlorobenzene	1.004	1.033	-2.9	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.417	-8.3	93	0.00
67 C	Ethyl Benzene	1.831	1.896	-3.5#	89	0.00
68 T	m/p-Xylenes	0.652	0.697	-6.9	90	0.00
69 T	o-Xylene	0.632	0.665	-5.2	89	0.00
70 T	Styrene	1.086	1.169	-7.6	92	0.00
71 P	Bromoform	0.313	0.338	-8.0	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.513	3.626	-3.2	91	0.00
74 T	N-amyl acetate	1.700	1.594	6.2	86	0.00
75 P	1,1,2,2-Tetrachloroethane	1.241	1.182	4.8	88	0.00
76 T	1,2,3-Trichloropropane	1.164	1.185	-1.8	90	0.00
77 T	Bromobenzene	0.871	0.872	-0.1	90	0.00
78 T	n-propylbenzene	4.149	4.286	-3.3	92	0.00
79 T	2-Chlorotoluene	2.517	2.542	-1.0	90	0.00
80 T	1,3,5-Trimethylbenzene	3.001	3.166	-5.5	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.382	0.378	1.0	89	0.00
82 T	4-Chlorotoluene	2.988	2.987	0.0	90	0.00
83 T	tert-Butylbenzene	2.922	2.943	-0.7	90	0.00
84 T	1,2,4-Trimethylbenzene	3.026	3.213	-6.2	94	0.00
85 T	sec-Butylbenzene	3.371	3.576	-6.1	92	0.00
86 T	p-Isopropyltoluene	3.068	3.274	-6.7	91	0.00
87 T	1,3-Dichlorobenzene	1.568	1.579	-0.7	93	0.00
88 T	1,4-Dichlorobenzene	1.622	1.645	-1.4	96	0.00
89 T	n-Butylbenzene	2.811	2.984	-6.2	93	0.00
90 T	Hexachloroethane	0.554	0.561	-1.3	92	0.00
91 T	1,2-Dichlorobenzene	1.525	1.573	-3.1	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.286	0.278	2.8	87	0.00
93 T	1,2,4-Trichlorobenzene	0.999	1.030	-3.1	94	0.00
94 T	Hexachlorobutadiene	0.451	0.469	-4.0	97	0.00
95 T	Naphthalene	3.259	3.268	-0.3	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
Data File : VX021700.D
Acq On : 01 Apr 2021 11:49
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 02 03:24:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.999	1.038	-3.9	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6