

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021727.D
 Acq On : 01 Apr 2021 22:32
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 03:34:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	50.000	42.743	14.5	75	0.00
3 P	Chloromethane	50.000	44.654	10.7	81	0.00
4 C	Vinyl Chloride	50.000	48.535	2.9#	84	0.00
5 T	Bromomethane	50.000	50.366	-0.7	91	0.00
6 T	Chloroethane	50.000	50.620	-1.2	89	0.00
7 T	Trichlorofluoromethane	50.000	51.218	-2.4	88	0.00
8 T	Diethyl Ether	50.000	53.389	-6.8	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.099	1.8	87	0.00
10 T	Methyl Iodide	50.000	50.662	-1.3	99	0.00
11 T	Tert butyl alcohol	250.000	249.601	0.2	91	0.00
12 CM	1,1-Dichloroethene	50.000	47.853	4.3#	88	0.00
13 T	Acrolein	250.000	227.849	8.9	83	0.00
14 T	Allyl chloride	50.000	49.203	1.6	87	0.00
15 T	Acrylonitrile	250.000	262.410	-5.0	91	0.00
16 T	Acetone	250.000	255.007	-2.0	92	0.00
17 T	Carbon Disulfide	50.000	39.963	20.1	73	0.00
18 T	Methyl Acetate	50.000	58.945	-17.9	100	0.00
19 T	Methyl tert-butyl Ether	50.000	53.325	-6.7	93	0.00
20 T	Methylene Chloride	50.000	49.034	1.9	89	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.135	1.7	87	0.00
22 T	Diisopropyl ether	50.000	53.734	-7.5	93	0.00
23 T	Vinyl Acetate	250.000	261.497	-4.6	89	0.00
24 P	1,1-Dichloroethane	50.000	51.914	-3.8	91	0.00
25 T	2-Butanone	250.000	259.989	-4.0	91	0.00
26 T	2,2-Dichloropropane	50.000	44.624	10.8	77	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.957	-1.9	89	0.00
28 T	Bromochloromethane	50.000	49.609	0.8	93	0.00
29 T	Tetrahydrofuran	250.000	261.548	-4.6	92	0.00
30 C	Chloroform	50.000	53.969	-7.9#	93	0.00
31 T	Cyclohexane	50.000	47.133	5.7	81	0.00
32 T	1,1,1-Trichloroethane	50.000	52.638	-5.3	91	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.872	-1.7	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	91	0.00
35 S	Dibromofluoromethane	50.000	49.428	1.1	95	0.00
36 T	1,1-Dichloropropene	50.000	48.257	3.5	87	0.00
37 T	Ethyl Acetate	50.000	49.014	2.0	91	0.00
38 T	Carbon Tetrachloride	50.000	51.084	-2.2	90	0.00
39 T	Methylcyclohexane	50.000	45.709	8.6	79	0.00
40 TM	Benzene	50.000	50.554	-1.1	89	0.00
41 T	Methacrylonitrile	50.000	49.802	0.4	91	0.00
42 TM	1,2-Dichloroethane	50.000	51.025	-2.0	91	0.00
43 T	Isopropyl Acetate	50.000	51.468	-2.9	91	0.00
44 TM	Trichloroethene	50.000	51.010	-2.0	92	0.00
45 C	1,2-Dichloropropane	50.000	51.647	-3.3#	91	0.00
46 T	Dibromomethane	50.000	50.666	-1.3	89	0.00
47 T	Bromodichloromethane	50.000	52.525	-5.0	92	0.00
48 T	Methyl methacrylate	50.000	52.037	-4.1	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
 Data File : VX021727.D
 Acq On : 01 Apr 2021 22:32
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 03:34:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	944.172	5.6	83	0.00
50 S	Toluene-d8	50.000	49.188	1.6	91	0.00
51 T	4-Methyl-2-Pentanone	250.000	270.161	-8.1	93	0.00
52 CM	Toluene	50.000	51.005	-2.0#	88	0.00
53 T	t-1,3-Dichloropropene	50.000	51.121	-2.2	88	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.305	-2.6	88	0.00
55 T	1,1,2-Trichloroethane	50.000	53.857	-7.7	94	0.00
56 T	Ethyl methacrylate	50.000	48.281	3.4	90	0.00
57 T	1,3-Dichloropropane	50.000	52.785	-5.6	92	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	262.426	-5.0	90	0.00
59 T	2-Hexanone	250.000	271.151	-8.5	94	0.00
60 T	Dibromochloromethane	50.000	54.728	-9.5	94	0.00
61 T	1,2-Dibromoethane	50.000	52.356	-4.7	90	0.00
62 S	4-Bromofluorobenzene	50.000	50.583	-1.2	95	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	91	0.00
64 T	Tetrachloroethene	50.000	54.182	-8.4	96	0.00
65 PM	Chlorobenzene	50.000	49.823	0.4	89	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.494	-5.0	92	0.00
67 C	Ethyl Benzene	50.000	50.540	-1.1#	89	0.00
68 T	m/p-Xylenes	100.000	103.029	-3.0	89	0.00
69 T	o-Xylene	50.000	52.394	-4.8	91	0.00
70 T	Styrene	50.000	52.028	-4.1	91	0.00
71 P	Bromoform	50.000	53.684	-7.4	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	92	0.00
73 T	Isopropylbenzene	50.000	52.414	-4.8	91	0.00
74 T	N-amyl acetate	50.000	52.629	-5.3	95	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.004	-2.0	93	0.00
76 T	1,2,3-Trichloropropane	50.000	53.155	-6.3	92	0.00
77 T	Bromobenzene	50.000	52.025	-4.0	93	0.00
78 T	n-propylbenzene	50.000	50.706	-1.4	89	0.00
79 T	2-Chlorotoluene	50.000	51.603	-3.2	91	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.201	-4.4	90	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.525	1.0	88	0.00
82 T	4-Chlorotoluene	50.000	51.330	-2.7	91	0.00
83 T	tert-Butylbenzene	50.000	50.653	-1.3	89	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.218	-4.4	91	0.00
85 T	sec-Butylbenzene	50.000	52.626	-5.3	90	0.00
86 T	p-Isopropyltoluene	50.000	52.175	-4.3	88	0.00
87 T	1,3-Dichlorobenzene	50.000	52.029	-4.1	95	0.00
88 T	1,4-Dichlorobenzene	50.000	49.329	1.3	92	0.00
89 T	n-Butylbenzene	50.000	49.931	0.1	86	0.00
90 T	Hexachloroethane	50.000	52.324	-4.6	93	0.00
91 T	1,2-Dichlorobenzene	50.000	52.209	-4.4	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.632	2.7	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.534	-1.1	91	0.00
94 T	Hexachlorobutadiene	50.000	46.919	6.2	86	0.00
95 T	Naphthalene	50.000	52.677	-5.4	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040121\
Data File : VX021727.D
Acq On : 01 Apr 2021 22:32
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 02 03:34:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.224	-0.4	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6