

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040121\  
 Data File : VX021702.D  
 Acq On : 01 Apr 2021 13:01  
 Operator : JC/MD  
 Sample : VX0401WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0401WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 4/2/2021 12:55:23 PM

Quant Time: Apr 02 03:25:32 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032421W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 24 12:25:26 2021  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response           | Conc  | Units  | Dev(Min) |
|------------------------------|----------------|------|--------------------|-------|--------|----------|
| -----                        |                |      |                    |       |        |          |
| Internal Standards           |                |      |                    |       |        |          |
| 1) Pentafluorobenzene        | 5.623          | 168  | 74816              | 50.00 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.823          | 114  | 118423             | 50.00 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.094         | 117  | 111614             | 50.00 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.059         | 152  | 53968              | 50.00 | ug/l   | 0.00     |
|                              |                |      |                    |       |        |          |
| System Monitoring Compounds  |                |      |                    |       |        |          |
| 33) 1,2-Dichloroethane-d4    | 6.023          | 65   | 59185              | 47.54 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = 95.08%  |       |        |          |
| 35) Dibromofluoromethane     | 5.458          | 113  | 42875              | 50.97 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = 101.94% |       |        |          |
| 50) Toluene-d8               | 8.694          | 98   | 150590             | 50.24 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = 100.48% |       |        |          |
| 62) 4-Bromofluorobenzene     | 11.118         | 95   | 56298              | 48.31 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = 96.62%  |       |        |          |
|                              |                |      |                    |       |        |          |
| Target Compounds             |                |      |                    |       | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.187          | 85   | 13355              | 16.24 | ug/l   | 96       |
| 3) Chloromethane             | 1.311          | 50   | 15958              | 16.91 | ug/l   | 97       |
| 4) Vinyl Chloride            | 1.393          | 62   | 14401              | 17.12 | ug/l   | 99       |
| 5) Bromomethane              | 1.622          | 94   | 7452               | 22.01 | ug/l   | 95       |
| 6) Chloroethane              | 1.705          | 64   | 9391               | 17.60 | ug/l   | 100      |
| 7) Trichlorofluoromethane    | 1.911          | 101  | 25201              | 18.42 | ug/l   | 96       |
| 8) Diethyl Ether             | 2.170          | 74   | 8382               | 17.66 | ug/l   | 76       |
| 9) 1,1,2-Trichlorotrifluo... | 2.364          | 101  | 14532              | 18.87 | ug/l   | 97       |
| 10) Methyl Iodide            | 2.487          | 142  | 12402              | 24.53 | ug/l   | 90       |
| 11) Tert butyl alcohol       | 3.011          | 59   | 16712              | 81.21 | ug/l   | 95       |
| 12) 1,1-Dichloroethene       | 2.352          | 96   | 12506              | 16.59 | ug/l   | 82       |
| 13) Acrolein                 | 2.270          | 56   | 8757               | 87.21 | ug/l   | 95       |
| 14) Allyl chloride           | 2.705          | 41   | 25583              | 16.72 | ug/l # | 89       |
| 15) Acrylonitrile            | 3.111          | 53   | 43057              | 88.04 | ug/l   | 98       |
| 16) Acetone                  | 2.417          | 43   | 39306              | 83.51 | ug/l # | 87       |
| 17) Carbon Disulfide         | 2.546          | 76   | 30268              | 14.30 | ug/l   | 100      |
| 18) Methyl Acetate           | 2.746          | 43   | 26799              | 19.71 | ug/l # | 89       |
| 19) Methyl tert-butyl Ether  | 3.164          | 73   | 49729              | 17.66 | ug/l   | 96       |
| 20) Methylene Chloride       | 2.828          | 84   | 15955              | 17.37 | ug/l # | 87       |
| 21) trans-1,2-Dichloroethene | 3.140          | 96   | 14492              | 17.50 | ug/l   | 96       |
| 22) Diisopropyl ether        | 3.823          | 45   | 53472              | 17.98 | ug/l # | 95       |
| 23) Vinyl Acetate            | 3.782          | 43   | 212783             | 84.61 | ug/l # | 93       |
| 24) 1,1-Dichloroethane       | 3.664          | 63   | 28966              | 18.07 | ug/l   | 97       |
| 25) 2-Butanone               | 4.629          | 43   | 60274              | 84.48 | ug/l   | 90       |
| 26) 2,2-Dichloropropane      | 4.546          | 77   | 24201              | 17.43 | ug/l   | 97       |
| 27) cis-1,2-Dichloroethene   | 4.558          | 96   | 16886              | 17.56 | ug/l   | 90       |
| 28) Bromochloromethane       | 4.976          | 49   | 14936              | 18.61 | ug/l   | 85       |
| 29) Tetrahydrofuran          | 5.082          | 42   | 38221              | 82.62 | ug/l   | 88       |
| 30) Chloroform               | 5.170          | 83   | 31575              | 18.66 | ug/l   | 95       |
| 31) Cyclohexane              | 5.541          | 56   | 22286              | 16.28 | ug/l   | 96       |
| 32) 1,1,1-Trichloroethane    | 5.458          | 97   | 27272              | 18.19 | ug/l   | 95       |
| 36) 1,1-Dichloropropene      | 5.764          | 75   | 20744              | 17.88 | ug/l   | 99       |
| 37) Ethyl Acetate            | 4.793          | 43   | 22982              | 17.13 | ug/l # | 93       |
| 38) Carbon Tetrachloride     | 5.747          | 117  | 24437              | 19.60 | ug/l   | 97       |
| 39) Methylcyclohexane        | 7.441          | 83   | 22509              | 17.50 | ug/l   | 97       |
| 40) Benzene                  | 6.105          | 78   | 61537              | 18.68 | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040121\  
 Data File : VX021702.D  
 Acq On : 01 Apr 2021 13:01  
 Operator : JC/MD  
 Sample : VX0401WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0401WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 4/2/2021 12:55:23 PM

Quant Time: Apr 02 03:25:32 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032421W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 24 12:25:26 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| 41) Methacrylonitrile         | 4.993  | 41   | 13402    | 17.22  | ug/l   | 89       |
| 42) 1,2-Dichloroethane        | 6.158  | 62   | 27319    | 19.12  | ug/l   | 93       |
| 43) Isopropyl Acetate         | 6.405  | 43   | 38495    | 17.67  | ug/l # | 90       |
| 44) Trichloroethene           | 7.182  | 130  | 16673    | 19.18  | ug/l   | 98       |
| 45) 1,2-Dichloropropane       | 7.482  | 63   | 17247    | 19.32  | ug/l   | 97       |
| 46) Dibromomethane            | 7.629  | 93   | 12484    | 19.19  | ug/l   | 92       |
| 47) Bromodichloromethane      | 7.870  | 83   | 25609    | 19.34  | ug/l   | 99       |
| 48) Methyl methacrylate       | 7.741  | 41   | 20140    | 17.33  | ug/l # | 82       |
| 49) 1,4-Dioxane               | 7.706  | 88   | 7280     | 349.91 | ug/l # | 84       |
| 51) 4-Methyl-2-Pentanone      | 8.617  | 43   | 125105   | 93.36  | ug/l   | 88       |
| 52) Toluene                   | 8.765  | 92   | 38832    | 19.08  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 9.018  | 75   | 23902    | 17.88  | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 8.412  | 75   | 26188    | 18.41  | ug/l # | 84       |
| 55) 1,1,2-Trichloroethane     | 9.194  | 97   | 16775    | 19.53  | ug/l   | 97       |
| 56) Ethyl methacrylate        | 9.159  | 69   | 22848    | 17.17  | ug/l # | 73       |
| 57) 1,3-Dichloropropane       | 9.353  | 76   | 28139    | 19.14  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 8.288  | 63   | 65859    | 90.72  | ug/l   | 95       |
| 59) 2-Hexanone                | 9.471  | 43   | 91866    | 91.30  | ug/l   | 84       |
| 60) Dibromochloromethane      | 9.565  | 129  | 18591    | 19.32  | ug/l   | 97       |
| 61) 1,2-Dibromoethane         | 9.653  | 107  | 17290    | 18.74  | ug/l   | 97       |
| 64) Tetrachloroethene         | 9.318  | 164  | 16451    | 22.00  | ug/l   | 96       |
| 65) Chlorobenzene             | 10.118 | 112  | 41608    | 18.56  | ug/l   | 96       |
| 66) 1,1,1,2-Tetrachloroethane | 10.200 | 131  | 16977    | 19.76  | ug/l   | 98       |
| 67) Ethyl Benzene             | 10.235 | 91   | 73296    | 17.93  | ug/l   | 100      |
| 68) m/p-Xylenes               | 10.341 | 106  | 55150    | 37.88  | ug/l   | 95       |
| 69) o-Xylene                  | 10.682 | 106  | 25368    | 17.99  | ug/l   | 91       |
| 70) Styrene                   | 10.694 | 104  | 44582    | 18.38  | ug/l   | 97       |
| 71) Bromoform                 | 10.841 | 173  | 13200    | 18.91  | ug/l # | 96       |
| 73) Isopropylbenzene          | 11.000 | 105  | 73071    | 19.27  | ug/l   | 100      |
| 74) N-amyl acetate            | 10.877 | 43   | 31304    | 17.06  | ug/l # | 89       |
| 75) 1,1,2,2-Tetrachloroethane | 11.253 | 83   | 24729    | 18.46  | ug/l   | 97       |
| 76) 1,2,3-Trichloropropane    | 11.277 | 75   | 24334m   | 19.37  | ug/l   |          |
| 77) Bromobenzene              | 11.235 | 156  | 18091    | 19.25  | ug/l   | 96       |
| 78) n-propylbenzene           | 11.341 | 91   | 83471    | 18.64  | ug/l   | 98       |
| 79) 2-Chlorotoluene           | 11.400 | 91   | 50861    | 18.72  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 11.488 | 105  | 60883    | 18.79  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 11.059 | 75   | 7083     | 17.17  | ug/l # | 82       |
| 82) 4-Chlorotoluene           | 11.494 | 91   | 59855    | 18.56  | ug/l   | 97       |
| 83) tert-Butylbenzene         | 11.753 | 119  | 57188    | 18.13  | ug/l   | 92       |
| 84) 1,2,4-Trimethylbenzene    | 11.788 | 105  | 62535    | 19.15  | ug/l   | 95       |
| 85) sec-Butylbenzene          | 11.930 | 105  | 68920    | 18.94  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 12.047 | 119  | 62549    | 18.89  | ug/l   | 96       |
| 87) 1,3-Dichlorobenzene       | 12.006 | 146  | 32562    | 19.24  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 12.083 | 146  | 32498    | 18.56  | ug/l   | 97       |
| 89) n-Butylbenzene            | 12.371 | 91   | 55192    | 18.19  | ug/l   | 95       |
| 90) Hexachloroethane          | 12.577 | 117  | 11313    | 18.93  | ug/l   | 97       |
| 91) 1,2-Dichlorobenzene       | 12.377 | 146  | 31334    | 19.04  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.983 | 75   | 5535     | 17.72  | ug/l   | 82       |
| 93) 1,2,4-Trichlorobenzene    | 13.630 | 180  | 19893    | 18.44  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.765 | 225  | 9334     | 19.17  | ug/l   | 97       |
| 95) Naphthalene               | 13.818 | 128  | 62024    | 17.63  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 14.001 | 180  | 20023    | 18.57  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040121\  
Data File : VX021702.D  
Acq On : 01 Apr 2021 13:01  
Operator : JC/MD  
Sample : VX0401WBS01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01

Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM


Quant Time: Apr 02 03:25:32 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032421W.M  
Quant Title : SW846 8260  
QLast Update : Wed Mar 24 12:25:26 2021  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


(#) = qualifier out of range (m) = manual integration (+) = signals summed

MMDadoda  
4/2/2021 12:55:23 PM




[illegible]




#1  
Pentafluorobenzene  
Concen: 50.00 ug/l  
RT: 5.623 min Scan# 771  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01





Tgt Ion: 168 Resp: 74816  
Ion Ratio Lower Upper  
168 100  
99 60.0 45.8 68.6



#2  
Dichlorodifluoromethane  
Concen: 16.24 ug/l  
RT: 1.187 min Scan# 17  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01



Tgt Ion: 85 Resp: 13355  
Ion Ratio Lower Upper  
85 100  
87 35.3 16.4 49.2









Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01

Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM

Abundance Scan 38 (1.311 min): VX021427.D\data.ms (-33)




#3  
Chloromethane  
Concen: 16.91 ug/l  
RT: 1.311 min Scan# 38  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 50 Resp: 15958  
Ion Ratio Lower Upper  
50 100  
52 31.6 26.4 39.6



#4  
Vinyl Chloride  
Concen: 17.12 ug/l  
RT: 1.393 min Scan# 52  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01


Tgt Ion: 62 Resp: 14401  
Ion Ratio Lower Upper  
62 100  
64 31.3 25.7 38.5



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01


Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM



#5  
Bromomethane  
Concen: 22.01 ug/l  
RT: 1.622 min Scan# 91  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Tgt Ion: 94 Resp: 7452  
Ion Ratio Lower Upper  
94 100  
96 100.4 76.4 114.6

Manual Integrations  
APPROVED



MMDadoda  
4/2/2021 12:55:23 PM




#6  
Chloroethane  
Concen: 17.60 ug/l  
RT: 1.705 min Scan# 105  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01



Tgt Ion: 64 Resp: 9391  
Ion Ratio Lower Upper  
64 100  
66 31.6 25.4 38.0









#7  
Trichlorofluoromethane  
Concen: 18.42 ug/l  
RT: 1.911 min Scan# 140  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion:101 Resp: 25201  
Ion Ratio Lower Upper  
101 100  
103 63.5 53.2 79.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM




#8  
Diethyl Ether  
Concen: 17.66 ug/l  
RT: 2.170 min Scan# 184  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 74 Resp: 8382  
Ion Ratio Lower Upper  
74 100  
45 111.2 44.5 133.5











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 18.87 ug/l  
RT: 2.364 min Scan# 217  
Delta R.T. 0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 101   | Resp: | 14532 |
| Ion      | Ratio | Lower | Upper |
| 101      | 100   |       |       |
| 85       | 45.9  | 34.9  | 52.3  |
| 151      | 75.1  | 61.4  | 92.2  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#10  
Methyl Iodide  
Concen: 24.53 ug/l  
RT: 2.487 min Scan# 238  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 142   | Resp: | 12402 |
| Ion      | Ratio | Lower | Upper |
| 142      | 100   |       |       |
| 127      | 48.8  | 33.3  | 49.9  |
| 141      | 16.1  | 11.4  | 17.0  |









#11  
Tert butyl alcohol  
Concen: 81.21 ug/l  
RT: 3.011 min Scan# 327  
Delta R.T. -0.012 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 59 Resp: 16712  
Ion Ratio Lower Upper  
59 100  
57 11.7 7.8 11.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#12  
1,1-Dichloroethene  
Concen: 16.59 ug/l  
RT: 2.352 min Scan# 215  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 96 Resp: 12506  
Ion Ratio Lower Upper  
96 100  
61 186.7 124.6 187.0  
98 66.9 50.6 76.0









#13  
Acrolein  
Concen: 87.21 ug/l  
RT: 2.270 min Scan# 201  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 56 Resp: 8757  
Ion Ratio Lower Upper  
56 100  
55 74.5 56.4 84.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#14  
Allyl chloride  
Concen: 16.72 ug/l  
RT: 2.705 min Scan# 275  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 41 Resp: 25583  
Ion Ratio Lower Upper  
41 100  
39 71.7 51.6 77.4  
76 30.0 30.5 45.7#









#15  
Acrylonitrile  
Concen: 88.04 ug/l  
RT: 3.111 min Scan# 344  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 53    | Resp: | 43057 |
| Ion      | Ratio | Lower | Upper |
| 53       | 100   |       |       |
| 52       | 82.2  | 65.3  | 97.9  |
| 51       | 38.0  | 28.2  | 42.2  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#16  
Acetone  
Concen: 83.51 ug/l  
RT: 2.417 min Scan# 226  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 43    | Resp: | 39306 |
| Ion      | Ratio | Lower | Upper |
| 43       | 100   |       |       |
| 58       | 27.1  | 27.8  | 41.8# |





#17  
Carbon Disulfide  
Concen: 14.30 ug/l  
RT: 2.546 min Scan# 248  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Tgt Ion: 76 Resp: 30268  
Ion Ratio Lower Upper  
76 100  
78 8.9 7.2 10.8

Manual Integrations  
APPROVED



MMDadoda  
4/2/2021 12:55:23 PM




#18  
Methyl Acetate  
Concen: 19.71 ug/l  
RT: 2.746 min Scan# 282  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01



Tgt Ion: 43 Resp: 26799  
Ion Ratio Lower Upper  
43 100  
74 21.8 21.9 32.9#





#19





Methyl tert-butyl Ether  
Concen: 17.66 ug/l  
RT: 3.164 min Scan# 353  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 73 Resp: 49729  
Ion Ratio Lower Upper  
73 100  
57 23.2 17.1 25.7

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01

Manual Integrations  
APPROVED




MMDadoda  
4/2/2021 12:55:23 PM




#20

Methylene Chloride  
Concen: 17.37 ug/l  
RT: 2.828 min Scan# 296  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01


Tgt Ion: 84 Resp: 15955  
Ion Ratio Lower Upper  
84 100  
49 130.8 88.6 132.8  
51 40.9 27.0 40.6#  
86 60.3 50.3 75.5





#21  
trans-1,2-Dichloroethene  
Concen: 17.50 ug/l  
RT: 3.140 min Scan# 349  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Tgt Ion: 96 Resp: 14492  
Ion Ratio Lower Upper  
96 100  
61 140.0 107.7 161.5  
98 61.6 50.4 75.6

Manual Integrations  
APPROVED



MMDadoda  
4/2/2021 12:55:23 PM




#22  
Diisopropyl ether  
Concen: 17.98 ug/l  
RT: 3.823 min Scan# 465  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01



Tgt Ion: 45 Resp: 53472  
Ion Ratio Lower Upper  
45 100  
43 70.7 58.0 87.0  
87 25.0 25.4 38.0#  
59 10.9 10.2 15.2






#23  
 Vinyl Acetate  
 Concen: 84.61 ug/l  
 RT: 3.782 min Scan# 458  
 Delta R.T. -0.000 min  
 Lab File: VX021702.D  
 Acq: 01 Apr 2021 13:01




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0401WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 4/2/2021 12:55:23 PM






Tgt Ion: 43 Resp: 212783  
 Ion Ratio Lower Upper  
 43 100  
 86 9.5 9.9 14.9#




#24  
 1,1-Dichloroethane  
 Concen: 18.07 ug/l  
 RT: 3.664 min Scan# 438  
 Delta R.T. -0.006 min  
 Lab File: VX021702.D  
 Acq: 01 Apr 2021 13:01


Tgt Ion: 63 Resp: 28966  
 Ion Ratio Lower Upper  
 63 100  
 98 6.5 3.8 11.2  
 100 3.7 2.5 7.5











#25  
2-Butanone  
Concen: 84.48 ug/l  
RT: 4.629 min Scan# 602  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01





Tgt Ion: 43 Resp: 60274  
Ion Ratio Lower Upper  
43 100  
72 24.1 23.7 35.5



#26  
2,2-Dichloropropane  
Concen: 17.43 ug/l  
RT: 4.546 min Scan# 588  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 77 Resp: 24201  
Ion Ratio Lower Upper  
77 100  
97 22.1 11.7 35.1



Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01

Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM



#27





cis-1,2-Dichloroethene  
Concen: 17.56 ug/l  
RT: 4.558 min Scan# 590  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 96 Resp: 16886  
Ion Ratio Lower Upper  
96 100  
61 161.1 0.0 285.6  
98 64.6 0.0 129.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01

Manual Integrations  
APPROVED




MMDadoda  
4/2/2021 12:55:23 PM




#28

Bromochloromethane  
Concen: 18.61 ug/l  
RT: 4.976 min Scan# 661  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 49 Resp: 14936  
Ion Ratio Lower Upper  
49 100  
129 1.3 0.0 4.0  
130 71.3 68.2 102.2









#29  
Tetrahydrofuran  
Concen: 82.62 ug/l  
RT: 5.082 min Scan# 679  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 42    | Resp: | 38221 |
| Ion      | Ratio | Lower | Upper |
| 42       | 100   |       |       |
| 72       | 42.2  | 41.7  | 62.5  |
| 71       | 40.8  | 38.4  | 57.6  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM




#30  
Chloroform  
Concen: 18.66 ug/l  
RT: 5.170 min Scan# 694  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 83    | Resp: | 31575 |
| Ion      | Ratio | Lower | Upper |
| 83       | 100   |       |       |
| 85       | 68.8  | 52.0  | 78.0  |









#31  
Cyclohexane  
Concen: 16.28 ug/l  
RT: 5.541 min Scan# 757  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01





Tgt Ion: 56 Resp: 22286  
Ion Ratio Lower Upper  
56 100  
69 34.5 25.3 37.9  
84 86.2 71.4 107.0



#32  
1,1,1-Trichloroethane  
Concen: 18.19 ug/l  
RT: 5.458 min Scan# 743  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 97 Resp: 27272  
Ion Ratio Lower Upper  
97 100  
99 65.5 51.5 77.3  
61 49.6 34.2 51.4



Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01

Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM



#33





1,2-Dichloroethane-d4  
Concen: 47.54 ug/l  
RT: 6.023 min Scan# 839  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 65 Resp: 59185  
Ion Ratio Lower Upper  
65 100  
67 50.0 0.0 105.4

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01

Manual Integrations  
APPROVED




MMDadoda  
4/2/2021 12:55:23 PM




#34


1,4-Difluorobenzene  
Concen: 50.00 ug/l  
RT: 6.823 min Scan# 975  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 114 Resp: 118423  
Ion Ratio Lower Upper  
114 100  
63 23.3 0.0 40.8  
88 17.6 0.0 32.8







#35  
Dibromofluoromethane  
Concen: 50.97 ug/l  
RT: 5.458 min Scan# 743  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 113 Resp: 42875  
Ion Ratio Lower Upper  
113 100  
111 100.5 81.9 122.9  
192 19.2 16.7 25.1




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01

Manual Integrations  
APPROVED



MMDadoda  
4/2/2021 12:55:23 PM




#36  
1,1-Dichloropropene  
Concen: 17.88 ug/l  
RT: 5.764 min Scan# 795  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 75 Resp: 20744  
Ion Ratio Lower Upper  
75 100  
110 34.8 18.1 54.4  
77 31.7 25.3 37.9







#37  
Ethyl Acetate  
Concen: 17.13 ug/l  
RT: 4.793 min Scan# 630  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01



Tgt Ion: 43 Resp: 22982


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 43  | 100   |       |       |
| 61  | 13.3  | 12.8  | 19.2  |
| 70  | 10.0  | 10.5  | 15.7# |




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01

Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM






#38  
Carbon Tetrachloride  
Concen: 19.60 ug/l  
RT: 5.747 min Scan# 792  
Delta R.T. -0.005 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01



Tgt Ion: 117 Resp: 24437


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 117 | 100   |       |       |
| 119 | 98.6  | 76.2  | 114.2 |
| 121 | 32.4  | 24.8  | 37.2  |





#39  
Methylcyclohexane  
Concen: 17.50 ug/l  
RT: 7.441 min Scan# 1080  
Delta R.T. 0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01



Tgt Ion: 83 Resp: 22509  
Ion Ratio Lower Upper  
83 100  
55 75.5 58.9 88.3  
98 45.5 38.1 57.1




Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM




#40  
Benzene  
Concen: 18.68 ug/l  
RT: 6.105 min Scan# 853  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01


Tgt Ion: 78 Resp: 61537  
Ion Ratio Lower Upper  
78 100  
77 24.5 18.9 28.3











#41  
Methacrylonitrile  
Concen: 17.22 ug/l  
RT: 4.993 min Scan# 664  
Delta R.T. -0.012 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01





Tgt Ion: 41 Resp: 13402  
Ion Ratio Lower Upper  
41 100  
39 59.6 42.5 63.7  
67 66.5 65.0 97.6  
52 32.4 25.3 37.9



#42  
1,2-Dichloroethane  
Concen: 19.12 ug/l  
RT: 6.158 min Scan# 862  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 62 Resp: 27319  
Ion Ratio Lower Upper  
62 100  
98 7.4 0.0 20.2



Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01

Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM







#43  
Isopropyl Acetate  
Concen: 17.67 ug/l  
RT: 6.405 min Scan# 904  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 43 Resp: 38495  
Ion Ratio Lower Upper  
43 100  
61 17.5 18.2 27.4#  
87 11.9 12.5 18.7#

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#44  
Trichloroethene  
Concen: 19.18 ug/l  
RT: 7.182 min Scan# 1036  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion:130 Resp: 16673  
Ion Ratio Lower Upper  
130 100  
95 101.3 0.0 198.0





#45





1,2-Dichloropropane  
Concen: 19.32 ug/l  
RT: 7.482 min Scan# 1087  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 63 Resp: 17247  
Ion Ratio Lower Upper  
63 100  
65 30.3 25.7 38.5

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01

Manual Integrations  
APPROVED




MMDadoda  
4/2/2021 12:55:23 PM




#46

Dibromomethane  
Concen: 19.19 ug/l  
RT: 7.629 min Scan# 1112  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 93 Resp: 12484  
Ion Ratio Lower Upper  
93 100  
95 81.2 66.8 100.2  
174 91.7 83.1 124.7









#47  
Bromodichloromethane  
Concen: 19.34 ug/l  
RT: 7.870 min Scan# 1153  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 83    | Resp: | 25609 |
| Ion      | Ratio | Lower | Upper |
| 83       | 100   |       |       |
| 85       | 63.3  | 51.4  | 77.0  |
| 127      | 8.2   | 6.8   | 10.2  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM




#48  
Methyl methacrylate  
Concen: 17.33 ug/l  
RT: 7.741 min Scan# 1131  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 41    | Resp: | 20140  |
| Ion      | Ratio | Lower | Upper  |
| 41       | 100   |       |        |
| 69       | 74.5  | 76.0  | 114.0# |
| 39       | 64.2  | 44.6  | 67.0   |









#49  
1,4-Dioxane  
Concen: 349.91 ug/l  
RT: 7.706 min Scan# 1125  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01





Tgt Ion: 88 Resp: 7280  
Ion Ratio Lower Upper  
88 100  
43 42.5 23.2 34.8#  
58 78.3 55.0 82.4



#50  
Toluene-d8  
Concen: 50.24 ug/l  
RT: 8.694 min Scan# 1293  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 98 Resp: 150590  
Ion Ratio Lower Upper  
98 100  
100 64.6 51.9 77.9



Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01

Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM







#51  
4-Methyl-2-Pentanone  
Concen: 93.36 ug/l  
RT: 8.617 min Scan# 1280  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 43 Resp: 125105  
Ion Ratio Lower Upper  
43 100  
58 36.6 35.4 53.0

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#52  
Toluene  
Concen: 19.08 ug/l  
RT: 8.765 min Scan# 1305  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 92 Resp: 38832  
Ion Ratio Lower Upper  
92 100  
91 172.8 137.1 205.7









#53  
t-1,3-Dichloropropene  
Concen: 17.88 ug/l  
RT: 9.018 min Scan# 1348  
Delta R.T. -0.005 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 75 Resp: 23902  
Ion Ratio Lower Upper  
75 100  
77 30.3 24.9 37.3

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#54  
cis-1,3-Dichloropropene  
Concen: 18.41 ug/l  
RT: 8.412 min Scan# 1245  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 75 Resp: 26188  
Ion Ratio Lower Upper  
75 100  
77 29.9 25.7 38.5  
39 59.4 35.1 52.7#






#55  
1,1,2-Trichloroethane  
Concen: 19.53 ug/l  
RT: 9.194 min Scan# 1378  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01

Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM






|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 97    | Resp: | 16775 |
| Ion      | Ratio | Lower | Upper |
| 97       | 100   |       |       |
| 83       | 90.6  | 69.3  | 103.9 |
| 85       | 57.7  | 44.3  | 66.5  |
| 99       | 64.0  | 50.3  | 75.5  |




#56  
Ethyl methacrylate  
Concen: 17.17 ug/l  
RT: 9.159 min Scan# 1372  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 69    | Resp: | 22848 |
| Ion      | Ratio | Lower | Upper |
| 69       | 100   |       |       |
| 41       | 76.6  | 46.8  | 70.2# |
| 39       | 48.1  | 24.7  | 37.1# |











#57  
1,3-Dichloropropane  
Concen: 19.14 ug/l  
RT: 9.353 min Scan# 1405  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 76 Resp: 28139  
Ion Ratio Lower Upper  
76 100  
78 32.5 26.2 39.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#58  
2-Chloroethyl Vinyl ether  
Concen: 90.72 ug/l  
RT: 8.288 min Scan# 1224  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 63 Resp: 65859  
Ion Ratio Lower Upper  
63 100  
106 26.9 23.5 35.3









#59  
2-Hexanone  
Concen: 91.30 ug/l  
RT: 9.471 min Scan# 1425  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 43 Resp: 91866  
Ion Ratio Lower Upper  
43 100  
58 49.2 30.9 92.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#60  
Dibromochloromethane  
Concen: 19.32 ug/l  
RT: 9.565 min Scan# 1441  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01


Tgt Ion: 129 Resp: 18591  
Ion Ratio Lower Upper  
129 100  
127 79.1 38.4 115.2





#61  
1,2-Dibromoethane  
Concen: 18.74 ug/l  
RT: 9.653 min Scan# 1456  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Tgt Ion:107 Resp: 17290  
Ion Ratio Lower Upper  
107 100  
109 91.6 75.8 113.6

Manual Integrations  
APPROVED



MMDadoda  
4/2/2021 12:55:23 PM




#62  
4-Bromofluorobenzene  
Concen: 48.31 ug/l  
RT: 11.118 min Scan# 1705  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 95 Resp: 56298  
Ion Ratio Lower Upper  
95 100  
174 77.9 0.0 154.6  
176 74.3 0.0 155.8





#63  
Chlorobenzene-d5  
Concen: 50.00 ug/l  
RT: 10.094 min Scan# 1531  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01




Tgt Ion:117 Resp: 111614  
Ion Ratio Lower Upper  
117 100  
82 58.1 47.4 71.2  
119 31.9 26.6 39.8

Manual Integrations  
APPROVED



MMDadoda  
4/2/2021 12:55:23 PM




#64  
Tetrachloroethene  
Concen: 22.00 ug/l  
RT: 9.318 min Scan# 1399  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion:164 Resp: 16451  
Ion Ratio Lower Upper  
164 100  
166 126.9 103.4 155.2  
129 98.1 70.8 106.2  
131 89.0 72.5 108.7







#65  
Chlorobenzene  
Concen: 18.56 ug/l  
RT: 10.118 min Scan# 1535  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion:112 Resp: 41608  
Ion Ratio Lower Upper  
112 100  
114 30.0 25.8 38.6




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01

Manual Integrations  
APPROVED



MMDadoda  
4/2/2021 12:55:23 PM




#66  
1,1,1,2-Tetrachloroethane  
Concen: 19.76 ug/l  
RT: 10.200 min Scan# 1549  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01



Tgt Ion:131 Resp: 16977  
Ion Ratio Lower Upper  
131 100  
133 96.3 47.7 143.1  
119 62.3 32.1 96.5









#67  
Ethyl Benzene  
Concen: 17.93 ug/l  
RT: 10.235 min Scan# 1555  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01

Tgt Ion: 91 Resp: 73296  
Ion Ratio Lower Upper  
91 100  
106 30.9 24.6 37.0




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#68  
m/p-Xylenes  
Concen: 37.88 ug/l  
RT: 10.341 min Scan# 1573  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01


Tgt Ion: 106 Resp: 55150  
Ion Ratio Lower Upper  
106 100  
91 210.8 162.8 244.2





#69  
o-Xylene  
Concen: 17.99 ug/l  
RT: 10.682 min Scan# 1631  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01



Tgt Ion:106 Resp: 25368  
Ion Ratio Lower Upper  
106 100  
91 230.0 108.1 324.2




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#70  
Styrene  
Concen: 18.38 ug/l  
RT: 10.694 min Scan# 1633  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion:104 Resp: 44582  
Ion Ratio Lower Upper  
104 100  
78 53.7 40.0 60.0  
103 54.9 43.4 65.0









#71  
Bromoform  
Concen: 18.91 ug/l  
RT: 10.841 min Scan# 1658  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 173     | 100   |       |       |
| 175     | 50.4  | 23.9  | 71.8  |
| 254     | 0.0   | 0.1   | 0.1#  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.00 ug/l  
RT: 12.059 min Scan# 1865  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 152     | 100   |       |       |
| 115     | 71.8  | 41.1  | 123.3 |
| 150     | 163.9 | 0.0   | 349.0 |











#73  
Isopropylbenzene  
Concen: 19.27 ug/l  
RT: 11.000 min Scan# 1685  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 105 Resp: 73071  
Ion Ratio Lower Upper  
105 100  
120 26.2 13.2 39.6

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM




#74  
N-ethyl acetate  
Concen: 17.06 ug/l  
RT: 10.877 min Scan# 1664  
Delta R.T. -0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 43 Resp: 31304  
Ion Ratio Lower Upper  
43 100  
70 38.7 39.0 58.6#  
55 24.9 21.5 32.3  
61 21.3 22.1 33.1#









#75  
1,1,2,2-Tetrachloroethane  
Concen: 18.46 ug/l  
RT: 11.253 min Scan# 1728  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01





Tgt Ion: 83 Resp: 24729  
Ion Ratio Lower Upper  
83 100  
131 9.9 5.1 15.2  
85 66.2 31.8 95.4



#76  
1,2,3-Trichloropropane  
Concen: 19.37 ug/l m  
RT: 11.277 min Scan# 1732  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 75 Resp: 24334  
Ion Ratio Lower Upper  
75 100  
77 37.1 19.1 57.5



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01

Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM



#77  
Bromobenzene  
Concen: 19.25 ug/l  
RT: 11.235 min Scan# 1725  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01





Tgt Ion: 156 Resp: 18091  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 156 | 100   |       |       |
| 77  | 170.6 | 82.8  | 248.6 |
| 158 | 94.1  | 49.1  | 147.4 |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01

Manual Integrations  
APPROVED




MMDadoda  
4/2/2021 12:55:23 PM




#78  
n-propylbenzene  
Concen: 18.64 ug/l  
RT: 11.341 min Scan# 1743  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 91 Resp: 83471  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 91  | 100   |       |       |
| 120 | 22.0  | 11.5  | 34.5  |









#79  
2-Chlorotoluene  
Concen: 18.72 ug/l  
RT: 11.400 min Scan# 1753  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 91 Resp: 50861  
Ion Ratio Lower Upper  
91 100  
126 33.8 16.6 49.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM




#80  
1,3,5-Trimethylbenzene  
Concen: 18.79 ug/l  
RT: 11.488 min Scan# 1768  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion: 105 Resp: 60883  
Ion Ratio Lower Upper  
105 100  
120 48.2 23.8 71.5









#81  
trans-1,4-Dichloro-2-butene  
Concen: 17.17 ug/l  
RT: 11.059 min Scan# 1695  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01





Tgt Ion: 75 Resp: 7083  
Ion Ratio Lower Upper  
75 100  
53 119.0 74.5 111.7#  
89 48.3 37.9 56.9



#82  
4-Chlorotoluene  
Concen: 18.56 ug/l  
RT: 11.494 min Scan# 1769  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 91 Resp: 59855  
Ion Ratio Lower Upper  
91 100  
126 27.9 14.9 44.5



Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01

Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM







#83  
tert-Butylbenzene  
Concen: 18.13 ug/l  
RT: 11.753 min Scan# 1813  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion:119 Resp: 57188  
Ion Ratio Lower Upper  
119 100  
91 64.2 27.9 83.7  
134 23.1 11.3 33.9

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01




Manual Integrations  
APPROVED


MMDadoda  
4/2/2021 12:55:23 PM



#84  
1,2,4-Trimethylbenzene  
Concen: 19.15 ug/l  
RT: 11.788 min Scan# 1819  
Delta R.T. -0.001 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01


Tgt Ion:105 Resp: 62535  
Ion Ratio Lower Upper  
105 100  
120 41.5 22.4 67.3





#85  
 sec-Butylbenzene  
 Concen: 18.94 ug/l  
 RT: 11.930 min Scan# 1843  
 Delta R.T. -0.000 min  
 Lab File: VX021702.D  
 Acq: 01 Apr 2021 13:01




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0401WBS01




Tgt Ion:105 Resp: 68920  
 Ion Ratio Lower Upper  
 105 100  
 134 19.6 10.0 29.8

Manual Integrations  
 APPROVED



MMDadoda  
 4/2/2021 12:55:23 PM




#86  
 p-Isopropyltoluene  
 Concen: 18.89 ug/l  
 RT: 12.047 min Scan# 1863  
 Delta R.T. 0.000 min  
 Lab File: VX021702.D  
 Acq: 01 Apr 2021 13:01



Tgt Ion:119 Resp: 62549  
 Ion Ratio Lower Upper  
 119 100  
 134 25.4 12.9 38.7  
 91 27.0 11.7 35.0









#87  
1,3-Dichlorobenzene  
Concen: 19.24 ug/l  
RT: 12.006 min Scan# 1856  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

Tgt Ion:146 Resp: 32562  
Ion Ratio Lower Upper  
146 100  
111 41.6 20.2 60.5  
148 63.1 31.9 95.7

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01




Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM




#88  
1,4-Dichlorobenzene  
Concen: 18.56 ug/l  
RT: 12.083 min Scan# 1869  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01


Tgt Ion:146 Resp: 32498  
Ion Ratio Lower Upper  
146 100  
111 43.2 19.6 58.8  
148 64.1 31.6 95.0











#89  
n-Butylbenzene  
Concen: 18.19 ug/l  
RT: 12.371 min Scan# 1918  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01





Tgt Ion: 91 Resp: 55192  
Ion Ratio Lower Upper  
91 100  
92 51.4 27.4 82.0  
134 23.5 13.2 39.5



#90  
Hexachloroethane  
Concen: 18.93 ug/l  
RT: 12.577 min Scan# 1953  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 117 Resp: 11313  
Ion Ratio Lower Upper  
117 100  
201 77.8 40.4 121.2



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01


Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM



#91  
1,2-Dichlorobenzene  
Concen: 19.04 ug/l  
RT: 12.377 min Scan# 1919  
Delta R.T. 0.006 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01




Tgt Ion:146 Resp: 31334  
Ion Ratio Lower Upper  
146 100  
111 44.6 20.8 62.2  
148 64.3 31.9 95.5

Manual Integrations  
APPROVED



MMDadoda  
4/2/2021 12:55:23 PM




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 17.72 ug/l  
RT: 12.983 min Scan# 2022  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion: 75 Resp: 5535  
Ion Ratio Lower Upper  
75 100  
155 72.4 45.3 135.9  
157 96.2 57.7 173.1









#93  
1,2,4-Trichlorobenzene  
Concen: 18.44 ug/l  
RT: 13.630 min Scan# 2132  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01





Tgt Ion:180 Resp: 19893  
Ion Ratio Lower Upper  
180 100  
182 96.5 47.7 143.1  
145 30.4 15.0 45.0



#94  
Hexachlorobutadiene  
Concen: 19.17 ug/l  
RT: 13.765 min Scan# 2155  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Tgt Ion:225 Resp: 9334  
Ion Ratio Lower Upper  
225 100  
223 63.1 31.1 93.2  
227 66.6 31.6 95.0



Instrument :  
MSVOA\_X  
ClientSampled :  
VX0401WBS01


Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM



#95  
Naphthalene  
Concen: 17.63 ug/l  
RT: 13.818 min Scan# 2164  
Delta R.T. 0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0401WBS01



|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 128   | Resp: | 62024 |
| Ion Ratio | Lower | Upper |       |
| 128       | 100   |       |       |
| 127       | 12.9  | 10.3  | 15.5  |
| 129       | 10.5  | 8.8   | 13.2  |

Manual Integrations  
APPROVED

MMDadoda  
4/2/2021 12:55:23 PM



#96  
1,2,3-Trichlorobenzene  
Concen: 18.57 ug/l  
RT: 14.001 min Scan# 2195  
Delta R.T. -0.000 min  
Lab File: VX021702.D  
Acq: 01 Apr 2021 13:01

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 180   | Resp: | 20023 |
| Ion Ratio | Lower | Upper |       |
| 180       | 100   |       |       |
| 182       | 95.0  | 47.6  | 142.9 |
| 145       | 33.2  | 16.1  | 48.3  |

