

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027879.D
 Acq On : 01 Apr 2022 14:26
 Operator : JC/MD
 Sample : N2074-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D3-20220320

Quant Time: Apr 02 05:06:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73298	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	129367	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116758	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51751	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59792	63.222	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 126.440%	
35) Dibromofluoromethane	5.391	113	44704	53.536	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.080%	
50) Toluene-d8	8.652	98	154602	52.621	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.240%	
62) 4-Bromofluorobenzene	11.085	95	62403	49.980	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.960%	
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	1945	4.647	ug/l	# 70
29) Tetrahydrofuran	5.013	42	20939	53.523	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027879.D
Acq On : 01 Apr 2022 14:26
Operator : JC/MD
Sample : N2074-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D3-20220320

Quant Time: Apr 02 05:06:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

