

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027880.D
 Acq On : 01 Apr 2022 14:49
 Operator : JC/MD
 Sample : N2074-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW1-7-20220321

Quant Time: Apr 02 05:07:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	78393	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	136290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56567	55.925	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	111.840%	
35) Dibromofluoromethane	5.391	113	45577	51.809	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	103.620%	
50) Toluene-d8	8.653	98	156601	50.594	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	101.180%	
62) 4-Bromofluorobenzene	11.079	95	63954	48.621	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	97.240%	
Target Compounds						
16) Acetone	2.379	43	760	1.698	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027880.D
Acq On : 01 Apr 2022 14:49
Operator : JC/MD
Sample : N2074-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW1-7-20220321

Quant Time: Apr 02 05:07:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

