

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027882.D
 Acq On : 01 Apr 2022 15:36
 Operator : JC/MD
 Sample : N2074-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D1-20220321

Quant Time: Apr 02 05:07:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76799	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	134183	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50698	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	47085	47.517	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.040%
35) Dibromofluoromethane	5.391	113	45904	53.000	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.000%
50) Toluene-d8	8.653	98	151476	49.707	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.420%
62) 4-Bromofluorobenzene	11.085	95	60873	47.005	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.020%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	796	1.107	ug/l	93
16) Acetone	2.380	43	816	1.861	ug/l	# 75
27) cis-1,2-Dichloroethene	4.495	96	248	0.287	ug/l	87
44) Trichloroethene	7.129	130	36583	40.720	ug/l	89
64) Tetrachloroethene	9.281	164	1829	2.298	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027882.D
Acq On : 01 Apr 2022 15:36
Operator : JC/MD
Sample : N2074-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D1-20220321

Quant Time: Apr 02 05:07:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

