

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027886.D
 Acq On : 01 Apr 2022 17:09
 Operator : JC/MD
 Sample : N2074-05DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D2-20220319DL

Quant Time: Apr 02 05:09:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	143990	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55686	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56610	53.380	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.760%	
35) Dibromofluoromethane	5.391	113	47448	51.051	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.100%	
50) Toluene-d8	8.653	98	163506	50.000	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.000%	
62) 4-Bromofluorobenzene	11.085	95	65409	47.068	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.140%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	28505	37.037	ug/l	97
12) 1,1-Dichloroethene	2.331	96	376	0.532	ug/l #	51
27) cis-1,2-Dichloroethene	4.495	96	941	1.018	ug/l #	71
44) Trichloroethene	7.129	130	26689	27.684	ug/l	97
64) Tetrachloroethene	9.281	164	2368	2.776	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027886.D
Acq On : 01 Apr 2022 17:09
Operator : JC/MD
Sample : N2074-05DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D2-20220319DL

Quant Time: Apr 02 05:09:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

