

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027887.D
 Acq On : 01 Apr 2022 17:32
 Operator : JC/MD
 Sample : N2074-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20220319DL

Quant Time: Apr 02 05:09:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79690	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	135207	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54840	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58579	56.971	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.940%
35) Dibromofluoromethane	5.391	113	46638	53.439	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.880%
50) Toluene-d8	8.653	98	162170	52.813	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.620%
62) 4-Bromofluorobenzene	11.085	95	65135	49.915	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.840%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	23071	30.918	ug/l	98
44) Trichloroethene	7.129	130	2133	2.356	ug/l	92
64) Tetrachloroethene	9.275	164	834	0.992	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027887.D
Acq On : 01 Apr 2022 17:32
Operator : JC/MD
Sample : N2074-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D3-20220319DL

Quant Time: Apr 02 05:09:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

