

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027889.D
 Acq On : 01 Apr 2022 18:18
 Operator : JC/MD
 Sample : N2074-08DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP04-20220319DL

Quant Time: Apr 02 05:10:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77253	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	136362	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124044	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59023	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62469	62.671	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	125.340%
35) Dibromofluoromethane	5.391	113	45786	52.019	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.040%
50) Toluene-d8	8.653	98	162396	52.439	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.880%
62) 4-Bromofluorobenzene	11.085	95	66996	50.906	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.820%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.330	101	413	0.571	ug/l	93
27) cis-1,2-Dichloroethene	4.489	96	649	0.747	ug/l	91
44) Trichloroethene	7.128	130	38052	41.678	ug/l	82
64) Tetrachloroethene	9.274	164	2240	2.663	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027889.D
Acq On : 01 Apr 2022 18:18
Operator : JC/MD
Sample : N2074-08DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP04-20220319DL

Quant Time: Apr 02 05:10:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

