

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027892.D
 Acq On : 01 Apr 2022 19:28
 Operator : JC/MD
 Sample : N2074-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-20220321

Quant Time: Apr 02 05:10:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	74950	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	132399	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124708	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60475	62.535	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	125.060%
35) Dibromofluoromethane	5.391	113	45306	53.014	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.020%
50) Toluene-d8	8.653	98	161582	53.737	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.480%
62) 4-Bromofluorobenzene	11.085	95	62168	48.652	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.300%
Target Compounds						
16) Acetone	2.380	43	907	2.119	ug/l	Qvalue 96
44) Trichloroethene	7.135	130	1716	1.936	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027892.D
Acq On : 01 Apr 2022 19:28
Operator : JC/MD
Sample : N2074-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D3-20220321

Quant Time: Apr 02 05:10:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

