

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027894.D
 Acq On : 01 Apr 2022 20:15
 Operator : JC/MD
 Sample : N2074-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW181S-20220321

Quant Time: Apr 02 05:11:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	80692	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	139693	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53402	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	52139	50.078	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.160%	
35) Dibromofluoromethane	5.391	113	46990	52.114	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.220%	
50) Toluene-d8	8.653	98	163436	51.516	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.040%	
62) 4-Bromofluorobenzene	11.085	95	63757	47.290	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.580%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	478	0.690	ug/l	# 77
16) Acetone	2.379	43	776	1.684	ug/l	# 87
27) cis-1,2-Dichloroethene	4.495	96	661	0.729	ug/l	89
30) Chloroform	5.099	83	810	0.508	ug/l	# 81
44) Trichloroethene	7.129	130	1159	1.239	ug/l	98
64) Tetrachloroethene	9.275	164	1152	1.395	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027894.D
Acq On : 01 Apr 2022 20:15
Operator : JC/MD
Sample : N2074-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW181S-20220321

Quant Time: Apr 02 05:11:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

