

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027904.D
 Acq On : 02 Apr 2022 00:53
 Operator : JC/MD
 Sample : N2172-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3R

Quant Time: Apr 02 07:32:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82869	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	138925	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57967	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56275	52.631	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.260%
35) Dibromofluoromethane	5.385	113	47488	52.957	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.920%
50) Toluene-d8	8.653	98	154024	48.818	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.640%
62) 4-Bromofluorobenzene	11.079	95	60969	45.472	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.940%
Target Compounds						
16) Acetone	2.374	43	692	1.462	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027904.D
Acq On : 02 Apr 2022 00:53
Operator : JC/MD
Sample : N2172-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3R

Quant Time: Apr 02 07:32:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

