

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027905.D
 Acq On : 02 Apr 2022 01:16
 Operator : JC/MD
 Sample : N2172-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5R

Quant Time: Apr 02 07:32:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81457	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	141821	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54526	51.879	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.760%	
35) Dibromofluoromethane	5.391	113	45706	49.929	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.860%	
50) Toluene-d8	8.653	98	165789	51.474	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.940%	
62) 4-Bromofluorobenzene	11.085	95	64887	47.406	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.820%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.172	85	1033	1.400	ug/l	81
16) Acetone	2.380	43	1672	3.595	ug/l	95
19) Methyl tert-butyl Ether	3.111	73	25498	9.563	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027905.D
Acq On : 02 Apr 2022 01:16
Operator : JC/MD
Sample : N2172-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5R

Quant Time: Apr 02 07:32:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

