

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027911.D
 Acq On : 02 Apr 2022 03:35
 Operator : JC/MD
 Sample : N2172-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-13

Quant Time: Apr 02 07:34:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79576	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	141676	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56353	54.885	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.760%	
35) Dibromofluoromethane	5.391	113	46436	50.778	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.560%	
50) Toluene-d8	8.653	98	159322	49.516	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.040%	
62) 4-Bromofluorobenzene	11.079	95	65622	47.992	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.980%	
Target Compounds						
16) Acetone	2.373	43	827	1.820	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027911.D
Acq On : 02 Apr 2022 03:35
Operator : JC/MD
Sample : N2172-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-13

Quant Time: Apr 02 07:34:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

