

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027916.D
 Acq On : 02 Apr 2022 05:31
 Operator : JC/MD
 Sample : N2204-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Apr 02 07:35:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	78946	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	138608	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55627	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58935	57.858	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	115.720%
35) Dibromofluoromethane	5.391	113	47567	53.166	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.340%
50) Toluene-d8	8.653	98	165902	52.703	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.400%
62) 4-Bromofluorobenzene	11.085	95	64832	48.464	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.920%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.953	59	287617	1469.954	ug/l	100
16) Acetone	2.380	43	830	1.841	ug/l	96
19) Methyl tert-butyl Ether	3.111	73	633445	245.128	ug/l	100
22) Diisopropyl ether	3.764	45	13344	5.347	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027916.D
Acq On : 02 Apr 2022 05:31
Operator : JC/MD
Sample : N2204-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Apr 02 07:35:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

