

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027918.D
 Acq On : 02 Apr 2022 06:18
 Operator : JC/MD
 Sample : N2204-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2R

Quant Time: Apr 02 07:35:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77762	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	141562	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128418	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54960	54.777	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.560%	
35) Dibromofluoromethane	5.391	113	47764	52.273	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.540%	
50) Toluene-d8	8.653	98	168875	52.528	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.060%	
62) 4-Bromofluorobenzene	11.079	95	65256	47.763	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.520%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.953	59	1429	7.415	ug/l	# 72
16) Acetone	2.380	43	1074	2.419	ug/l	# 81
19) Methyl tert-butyl Ether	3.117	73	13598	5.342	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027918.D
Acq On : 02 Apr 2022 06:18
Operator : JC/MD
Sample : N2204-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2R

Quant Time: Apr 02 07:35:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

