

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X040122\
 Data File : VX027896.D
 Acq On : 01 Apr 2022 21:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 05:11:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	50.000	48.654	2.7	72	0.00
3 P	Chloromethane	50.000	48.356	3.3	80	0.00
4 C	Vinyl Chloride	50.000	50.336	-0.7#	81	0.00
5 T	Bromomethane	50.000	43.657	12.7	78	0.00
6 T	Chloroethane	50.000	45.289	9.4	79	0.00
7 T	Trichlorofluoromethane	50.000	50.984	-2.0	83	0.00
8 T	Diethyl Ether	50.000	54.277	-8.6	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.995	-2.0	85	0.00
10 T	Methyl Iodide	50.000	49.548	0.9	80	0.00
11 T	Tert butyl alcohol	250.000	267.706	-7.1	86	0.00
12 CM	1,1-Dichloroethene	50.000	48.033	3.9#	84	0.00
13 T	Acrolein	250.000	329.861	-31.9#	107	0.00
14 T	Allyl chloride	50.000	47.674	4.7	81	0.00
15 T	Acrylonitrile	250.000	283.388	-13.4	90	0.00
16 T	Acetone	250.000	244.766	2.1	81	0.00
17 T	Carbon Disulfide	50.000	38.418	23.2	69	0.00
18 T	Methyl Acetate	50.000	50.026	-0.1	84	0.00
19 T	Methyl tert-butyl Ether	50.000	54.372	-8.7	88	0.00
20 T	Methylene Chloride	50.000	49.353	1.3	85	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.337	5.3	82	0.00
22 T	Diisopropyl ether	50.000	56.846	-13.7	103	0.00
23 T	Vinyl Acetate	250.000	290.215	-16.1	104	0.00
24 P	1,1-Dichloroethane	50.000	54.001	-8.0	97	0.00
25 T	2-Butanone	250.000	289.458	-15.8	105	0.00
26 T	2,2-Dichloropropane	50.000	48.322	3.4	85	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.410	-4.8	88	0.00
28 T	Bromochloromethane	50.000	53.810	-7.6	97	0.00
29 T	Tetrahydrofuran	250.000	271.574	-8.6	98	0.00
30 C	Chloroform	50.000	52.859	-5.7#	93	0.00
31 T	Cyclohexane	50.000	48.881	2.2	85	0.00
32 T	1,1,1-Trichloroethane	50.000	51.921	-3.8	90	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.887	0.2	84	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	90	0.00
35 S	Dibromofluoromethane	50.000	51.717	-3.4	94	0.00
36 T	1,1-Dichloropropene	50.000	48.578	2.8	85	0.00
37 T	Ethyl Acetate	50.000	53.178	-6.4	102	0.00
38 T	Carbon Tetrachloride	50.000	51.039	-2.1	87	0.00
39 T	Methylcyclohexane	50.000	44.461	11.1	78	0.00
40 TM	Benzene	50.000	47.422	5.2	80	0.00
41 T	Methacrylonitrile	50.000	52.459	-4.9	95	0.00
42 TM	1,2-Dichloroethane	50.000	47.697	4.6	76	0.00
43 T	Isopropyl Acetate	50.000	47.389	5.2	73	0.00
44 TM	Trichloroethene	50.000	49.451	1.1	87	0.00
45 C	1,2-Dichloropropane	50.000	51.678	-3.4#	87	0.00
46 T	Dibromomethane	50.000	50.289	-0.6	85	0.00
47 T	Bromodichloromethane	50.000	50.932	-1.9	86	0.00
48 T	Methyl methacrylate	50.000	52.039	-4.1	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027896.D
 Acq On : 01 Apr 2022 21:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 05:11:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1017.322	-1.7	83	0.00
50 S	Toluene-d8	50.000	50.147	-0.3	90	0.00
51 T	4-Methyl-2-Pentanone	250.000	271.003	-8.4	89	0.00
52 CM	Toluene	50.000	49.366	1.3#	84	0.00
53 T	t-1,3-Dichloropropene	50.000	49.968	0.1	82	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.795	0.4	82	0.00
55 T	1,1,2-Trichloroethane	50.000	51.311	-2.6	88	0.00
56 T	Ethyl methacrylate	50.000	52.019	-4.0	85	0.00
57 T	1,3-Dichloropropane	50.000	51.063	-2.1	86	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	270.018	-8.0	87	0.00
59 T	2-Hexanone	250.000	274.586	-9.8	89	0.00
60 T	Dibromochloromethane	50.000	52.790	-5.6	88	0.00
61 T	1,2-Dibromoethane	50.000	51.165	-2.3	87	0.00
62 S	4-Bromofluorobenzene	50.000	51.333	-2.7	91	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	88	0.00
64 T	Tetrachloroethene	50.000	49.803	0.4	86	0.00
65 PM	Chlorobenzene	50.000	49.817	0.4	87	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.524	-5.0	88	0.00
67 C	Ethyl Benzene	50.000	50.265	-0.5#	85	0.00
68 T	m/p-Xylenes	100.000	100.737	-0.7	85	0.00
69 T	o-Xylene	50.000	51.139	-2.3	85	0.00
70 T	Styrene	50.000	52.806	-5.6	88	0.00
71 P	Bromoform	50.000	50.436	-0.9	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	92	0.00
73 T	Isopropylbenzene	50.000	47.798	4.4	86	0.00
74 T	N-ethyl acetate	50.000	50.593	-1.2	85	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.728	-3.5	91	0.00
76 T	1,2,3-Trichloropropane	50.000	52.114	-4.2	89	0.00
77 T	Bromobenzene	50.000	49.757	0.5	88	0.00
78 T	n-propylbenzene	50.000	48.478	3.0	88	0.00
79 T	2-Chlorotoluene	50.000	47.405	5.2	87	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.145	3.7	86	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.173	7.7	84	0.00
82 T	4-Chlorotoluene	50.000	47.916	4.2	86	0.00
83 T	tert-Butylbenzene	50.000	48.846	2.3	89	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.486	1.0	88	0.00
85 T	sec-Butylbenzene	50.000	49.839	0.3	88	0.00
86 T	p-Isopropyltoluene	50.000	49.632	0.7	87	0.00
87 T	1,3-Dichlorobenzene	50.000	48.808	2.4	90	0.00
88 T	1,4-Dichlorobenzene	50.000	48.071	3.9	90	0.00
89 T	n-Butylbenzene	50.000	49.343	1.3	86	0.00
90 T	Hexachloroethane	50.000	49.946	0.1	90	0.00
91 T	1,2-Dichlorobenzene	50.000	49.372	1.3	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	53.037	-6.1	91	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.190	1.6	91	0.00
94 T	Hexachlorobutadiene	50.000	50.341	-0.7	91	0.00
95 T	Naphthalene	50.000	51.183	-2.4	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027896.D
Acq On : 01 Apr 2022 21:01
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 02 05:11:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.094	-0.2	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6