

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX040122\
 Data File : VX027898.D
 Acq On : 01 Apr 2022 22:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 05:12:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	50.000	47.612	4.8	73	0.00
3 P	Chloromethane	50.000	46.480	7.0	79	0.00
4 C	Vinyl Chloride	50.000	49.168	1.7#	81	0.00
5 T	Bromomethane	50.000	42.687	14.6	78	0.00
6 T	Chloroethane	50.000	45.929	8.1	82	0.00
7 T	Trichlorofluoromethane	50.000	50.337	-0.7	84	0.00
8 T	Diethyl Ether	50.000	51.619	-3.2	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.436	1.1	84	0.00
10 T	Methyl Iodide	50.000	48.155	3.7	79	0.00
11 T	Tert butyl alcohol	250.000	271.090	-8.4	89	0.00
12 CM	1,1-Dichloroethene	50.000	46.495	7.0#	83	0.00
13 T	Acrolein	250.000	328.732	-31.5#	109	0.00
14 T	Allyl chloride	50.000	47.428	5.1	82	0.00
15 T	Acrylonitrile	250.000	286.181	-14.5	93	0.00
16 T	Acetone	250.000	250.363	-0.1	85	0.00
17 T	Carbon Disulfide	50.000	37.599	24.8	69	0.00
18 T	Methyl Acetate	50.000	50.026	-0.1	86	0.00
19 T	Methyl tert-butyl Ether	50.000	53.568	-7.1	89	0.00
20 T	Methylene Chloride	50.000	48.313	3.4	85	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.266	9.5	81	0.00
22 T	Diisopropyl ether	50.000	55.167	-10.3	103	0.00
23 T	Vinyl Acetate	250.000	284.194	-13.7	104	0.00
24 P	1,1-Dichloroethane	50.000	52.996	-6.0	98	0.00
25 T	2-Butanone	250.000	281.869	-12.7	105	0.00
26 T	2,2-Dichloropropane	50.000	47.328	5.3	86	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.846	-1.7	87	0.00
28 T	Bromochloromethane	50.000	50.196	-0.4	93	0.00
29 T	Tetrahydrofuran	250.000	271.920	-8.8	100	0.00
30 C	Chloroform	50.000	51.768	-3.5#	93	0.00
31 T	Cyclohexane	50.000	48.578	2.8	86	0.00
32 T	1,1,1-Trichloroethane	50.000	50.217	-0.4	89	0.00
33 S	1,2-Dichloroethane-d4	50.000	52.478	-5.0	90	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	91	0.00
35 S	Dibromofluoromethane	50.000	51.358	-2.7	95	0.00
36 T	1,1-Dichloropropene	50.000	49.061	1.9	87	0.00
37 T	Ethyl Acetate	50.000	49.126	1.7	96	0.00
38 T	Carbon Tetrachloride	50.000	51.180	-2.4	88	0.00
39 T	Methylcyclohexane	50.000	44.824	10.4	80	0.00
40 TM	Benzene	50.000	49.593	0.8	85	0.00
41 T	Methacrylonitrile	50.000	52.455	-4.9	96	0.00
42 TM	1,2-Dichloroethane	50.000	50.026	-0.1	81	0.00
43 T	Isopropyl Acetate	50.000	47.008	6.0	74	0.00
44 TM	Trichloroethene	50.000	50.732	-1.5	91	0.00
45 C	1,2-Dichloropropane	50.000	51.225	-2.5#	87	0.00
46 T	Dibromomethane	50.000	49.505	1.0	85	0.00
47 T	Bromodichloromethane	50.000	49.929	0.1	86	0.00
48 T	Methyl methacrylate	50.000	50.644	-1.3	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027898.D
 Acq On : 01 Apr 2022 22:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 05:12:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1010.820	-1.1	84	0.00
50 S	Toluene-d8	50.000	49.241	1.5	90	0.00
51 T	4-Methyl-2-Pentanone	250.000	269.632	-7.9	89	0.00
52 CM	Toluene	50.000	49.922	0.2#	86	0.00
53 T	t-1,3-Dichloropropene	50.000	49.227	1.5	82	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.159	-0.3	84	0.00
55 T	1,1,2-Trichloroethane	50.000	50.737	-1.5	89	0.00
56 T	Ethyl methacrylate	50.000	52.215	-4.4	87	0.00
57 T	1,3-Dichloropropane	50.000	50.757	-1.5	87	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	263.313	-5.3	87	0.00
59 T	2-Hexanone	250.000	273.296	-9.3	90	0.00
60 T	Dibromochloromethane	50.000	51.022	-2.0	86	0.00
61 T	1,2-Dibromoethane	50.000	49.711	0.6	85	0.00
62 S	4-Bromofluorobenzene	50.000	51.339	-2.7	92	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	89	0.00
64 T	Tetrachloroethene	50.000	49.671	0.7	86	0.00
65 PM	Chlorobenzene	50.000	49.486	1.0	87	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.412	-2.8	87	0.00
67 C	Ethyl Benzene	50.000	50.018	-0.0#	85	0.00
68 T	m/p-Xylenes	100.000	99.915	0.1	85	0.00
69 T	o-Xylene	50.000	51.299	-2.6	86	0.00
70 T	Styrene	50.000	52.871	-5.7	89	0.00
71 P	Bromoform	50.000	49.325	1.3	89	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	47.336	5.3	88	0.00
74 T	N-amyl acetate	50.000	48.558	2.9	85	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.885	-1.8	93	0.00
76 T	1,2,3-Trichloropropane	50.000	51.178	-2.4	90	0.00
77 T	Bromobenzene	50.000	47.704	4.6	87	0.00
78 T	n-propylbenzene	50.000	46.607	6.8	87	0.00
79 T	2-Chlorotoluene	50.000	46.307	7.4	88	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.229	5.5	87	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.303	9.4	85	0.00
82 T	4-Chlorotoluene	50.000	46.757	6.5	87	0.00
83 T	tert-Butylbenzene	50.000	47.227	5.5	88	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.550	4.9	88	0.00
85 T	sec-Butylbenzene	50.000	48.421	3.2	88	0.00
86 T	p-Isopropyltoluene	50.000	48.867	2.3	88	0.00
87 T	1,3-Dichlorobenzene	50.000	47.730	4.5	91	0.00
88 T	1,4-Dichlorobenzene	50.000	46.315	7.4	89	0.00
89 T	n-Butylbenzene	50.000	48.291	3.4	87	0.00
90 T	Hexachloroethane	50.000	48.041	3.9	89	0.00
91 T	1,2-Dichlorobenzene	50.000	48.464	3.1	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.796	-5.6	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.316	5.4	90	0.00
94 T	Hexachlorobutadiene	50.000	48.420	3.2	91	0.00
95 T	Naphthalene	50.000	49.809	0.4	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027898.D
Acq On : 01 Apr 2022 22:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 02 05:12:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.657	2.7	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6