

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX040122\
 Data File : VX027921.D
 Acq On : 02 Apr 2022 07:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 08:24:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	50.000	47.274	5.5	66	0.00
3 P	Chloromethane	50.000	42.564	14.9	66	0.00
4 C	Vinyl Chloride	50.000	45.301	9.4#	69	0.00
5 T	Bromomethane	50.000	31.618	36.8#	53	0.00
6 T	Chloroethane	50.000	44.749	10.5	73	0.00
7 T	Trichlorofluoromethane	50.000	48.210	3.6	74	0.00
8 T	Diethyl Ether	50.000	51.338	-2.7	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.927	2.1	77	0.00
10 T	Methyl Iodide	50.000	32.738	34.5#	50	0.00
11 T	Tert butyl alcohol	250.000	244.399	2.2	74	0.00
12 CM	1,1-Dichloroethene	50.000	45.820	8.4#	76	0.00
13 T	Acrolein	250.000	279.984	-12.0	86	0.00
14 T	Allyl chloride	50.000	47.693	4.6	76	0.00
15 T	Acrylonitrile	250.000	268.769	-7.5	81	0.00
16 T	Acetone	250.000	249.021	0.4	78	0.00
17 T	Carbon Disulfide	50.000	37.680	24.6	64	0.00
18 T	Methyl Acetate	50.000	50.610	-1.2	80	0.00
19 T	Methyl tert-butyl Ether	50.000	52.219	-4.4	80	0.00
20 T	Methylene Chloride	50.000	48.031	3.9	78	0.00
21 T	trans-1,2-Dichloroethene	50.000	43.749	12.5	72	0.00
22 T	Diisopropyl ether	50.000	53.969	-7.9	93	0.00
23 T	Vinyl Acetate	250.000	270.752	-8.3	91	0.00
24 P	1,1-Dichloroethane	50.000	51.440	-2.9	87	0.00
25 T	2-Butanone	250.000	267.442	-7.0	91	0.00
26 T	2,2-Dichloropropane	50.000	38.963	22.1	65	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.313	3.4	76	0.00
28 T	Bromochloromethane	50.000	51.144	-2.3	87	0.01
29 T	Tetrahydrofuran	250.000	264.908	-6.0	90	0.00
30 C	Chloroform	50.000	52.633	-5.3#	87	0.00
31 T	Cyclohexane	50.000	45.980	8.0	75	0.00
32 T	1,1,1-Trichloroethane	50.000	50.849	-1.7	83	0.00
33 S	1,2-Dichloroethane-d4	50.000	54.340	-8.7	86	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	91	0.00
35 S	Dibromofluoromethane	50.000	49.236	1.5	91	0.00
36 T	1,1-Dichloropropene	50.000	45.082	9.8	80	0.00
37 T	Ethyl Acetate	50.000	46.154	7.7	90	0.00
38 T	Carbon Tetrachloride	50.000	46.155	7.7	79	0.00
39 T	Methylcyclohexane	50.000	38.463	23.1	68	0.00
40 TM	Benzene	50.000	44.162	11.7	76	0.00
41 T	Methacrylonitrile	50.000	49.142	1.7	90	0.00
42 TM	1,2-Dichloroethane	50.000	44.254	11.5	71	0.00
43 T	Isopropyl Acetate	50.000	42.399	15.2	66	0.00
44 TM	Trichloroethene	50.000	44.079	11.8	79	0.00
45 C	1,2-Dichloropropane	50.000	44.596	10.8#	76	0.00
46 T	Dibromomethane	50.000	43.878	12.2	75	0.00
47 T	Bromodichloromethane	50.000	44.660	10.7	76	0.00
48 T	Methyl methacrylate	50.000	40.238	19.5	66	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
 Data File : VX027921.D
 Acq On : 02 Apr 2022 07:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 08:24:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	915.110	8.5	76	0.00
50 S	Toluene-d8	50.000	47.408	5.2	86	0.00
51 T	4-Methyl-2-Pentanone	250.000	230.254	7.9	76	0.00
52 CM	Toluene	50.000	44.950	10.1#	77	0.00
53 T	t-1,3-Dichloropropene	50.000	42.398	15.2	70	0.00
54 T	cis-1,3-Dichloropropene	50.000	42.722	14.6	71	0.00
55 T	1,1,2-Trichloroethane	50.000	46.433	7.1	81	0.00
56 T	Ethyl methacrylate	50.000	45.706	8.6	75	0.00
57 T	1,3-Dichloropropane	50.000	45.183	9.6	77	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	227.094	9.2	74	0.00
59 T	2-Hexanone	250.000	232.696	6.9	76	0.00
60 T	Dibromochloromethane	50.000	46.811	6.4	79	0.00
61 T	1,2-Dibromoethane	50.000	44.603	10.8	76	0.00
62 S	4-Bromofluorobenzene	50.000	47.864	4.3	85	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	89	0.00
64 T	Tetrachloroethene	50.000	44.014	12.0	77	0.00
65 PM	Chlorobenzene	50.000	43.944	12.1	77	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	46.117	7.8	78	0.00
67 C	Ethyl Benzene	50.000	44.396	11.2#	76	0.00
68 T	m/p-Xylenes	100.000	87.928	12.1	75	0.00
69 T	o-Xylene	50.000	45.367	9.3	76	0.00
70 T	Styrene	50.000	45.721	8.6	77	0.00
71 P	Bromoform	50.000	43.216	13.6	78	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	93	0.00
73 T	Isopropylbenzene	50.000	42.233	15.5	77	0.00
74 T	N-amyl acetate	50.000	42.755	14.5	73	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	46.965	6.1	84	0.00
76 T	1,2,3-Trichloropropane	50.000	45.218	9.6	78	0.00
77 T	Bromobenzene	50.000	43.840	12.3	78	0.00
78 T	n-propylbenzene	50.000	41.419	17.2	76	0.00
79 T	2-Chlorotoluene	50.000	41.879	16.2	78	0.00
80 T	1,3,5-Trimethylbenzene	50.000	42.973	14.1	78	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	35.038	29.9#	63	0.00
82 T	4-Chlorotoluene	50.000	41.709	16.6	76	0.00
83 T	tert-Butylbenzene	50.000	42.704	14.6	78	0.00
84 T	1,2,4-Trimethylbenzene	50.000	42.588	14.8	77	0.00
85 T	sec-Butylbenzene	50.000	43.094	13.8	77	0.00
86 T	p-Isopropyltoluene	50.000	43.148	13.7	76	0.00
87 T	1,3-Dichlorobenzene	50.000	43.120	13.8	80	0.00
88 T	1,4-Dichlorobenzene	50.000	42.379	15.2	80	0.00
89 T	n-Butylbenzene	50.000	41.850	16.3	74	0.00
90 T	Hexachloroethane	50.000	42.736	14.5	77	0.00
91 T	1,2-Dichlorobenzene	50.000	45.119	9.8	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	45.422	9.2	78	0.00
93 T	1,2,4-Trichlorobenzene	50.000	43.387	13.2	81	0.00
94 T	Hexachlorobutadiene	50.000	44.388	11.2	81	0.00
95 T	Naphthalene	50.000	45.336	9.3	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027921.D
Acq On : 02 Apr 2022 07:27
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 02 08:24:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	44.253	11.5	79	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6