


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040122\
Data File : VX027874.D
Acq On : 01 Apr 2022 09:42
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Title : SW846 8260
Last Update : Wed Mar 30 04:39:27 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	3885	PASS
75	95	30	60	51.2	11311	PASS
95	95	100	100	100.0	22077	PASS
96	95	5	9	6.5	1443	PASS
173	174	0.00	2	0.6	92	PASS
174	95	50	100	72.0	15891	PASS
175	174	5	9	7.7	1221	PASS
176	174	95	101	96.9	15395	PASS
177	176	5	9	6.6	1013	PASS