

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040219\
 Data File : VX008586.D
 Acq On : 02 Apr 2019 11:07
 Operator : JC/SP
 Sample : VX0402WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0402WBL01

Quant Time: Apr 02 15:24:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X032719W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Mar 27 12:01:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	36324	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	212898	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	182690	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.06	65	92742	30.52	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.73%
60) 4-Bromofluorobenzene	11.13	95	89302	28.06	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	93.53%
63) Toluene-d8	8.71	98	264264	30.47	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.57%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040219\
Data File : VX008586.D
Acq On : 02 Apr 2019 11:07
Operator : JC/SP
Sample : VX0402WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0402WBL01

Quant Time: Apr 02 15:24:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X032719W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Mar 27 12:01:15 2019
Response via : Initial Calibration

