

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040219\
 Data File : VX008587.D
 Acq On : 02 Apr 2019 11:40
 Operator : JC/SP
 Sample : K2208-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EFFL-2

Quant Time: Apr 02 15:25:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X032719W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Mar 27 12:01:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	36673	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	211788	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	183662	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.06	65	91494	29.82	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	99.40%
60) 4-Bromofluorobenzene	11.13	95	89693	28.03	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	93.43%
63) Toluene-d8	8.71	98	264190	30.30	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.00%

Target Compounds

					Ovalue
15) Acetone	2.45	58	1632	3.041	ug/l 90
16) Carbon Disulfide	2.57	76	5382	0.745	ug/l # 94
62) Toluene	8.78	91	13369	1.202	ug/l 99
67) m/p-Xylenes	10.35	106	3631	0.807	ug/l 87
68) o-Xylene	10.69	106	2120	0.481	ug/l 95
70) Isopropylbenzene	11.02	105	3132	0.270	ug/l 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040219\
Data File : VX008587.D
Acq On : 02 Apr 2019 11:40
Operator : JC/SP
Sample : K2208-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EFFL-2

Quant Time: Apr 02 15:25:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X032719W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Mar 27 12:01:15 2019
Response via : Initial Calibration

