

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040219\
 Data File : VX008590.D
 Acq On : 02 Apr 2019 12:50
 Operator : JC/SP
 Sample : K2187-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-TRACK-89-90

Quant Time: Apr 03 11:28:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X032719W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Mar 27 12:01:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	34324	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	198615	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	169053	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.06	65	88060	30.67	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.23%
60) 4-Bromofluorobenzene	11.13	95	81584	27.70	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.33%
63) Toluene-d8	8.71	98	246179	30.68	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	102.27%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040219\
Data File : VX008590.D
Acq On : 02 Apr 2019 12:50
Operator : JC/SP
Sample : K2187-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-TRACK-89-90

Quant Time: Apr 03 11:28:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X032719W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Mar 27 12:01:15 2019
Response via : Initial Calibration

