

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040219\
 Data File : VX008598.D
 Acq On : 02 Apr 2019 18:15
 Operator : JC/SP
 Sample : K2210-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Apr 03 04:25:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	210923	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	350611	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	302479	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	125698	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	146309	45.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.52%	
35) Dibromofluoromethane	5.49	113	107360	46.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.92%	
50) Toluene-d8	8.71	98	437761	47.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.78%	
62) 4-Bromofluorobenzene	11.13	95	148746	43.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.82%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040219\
Data File : VX008598.D
Acq On : 02 Apr 2019 18:15
Operator : JC/SP
Sample : K2210-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Apr 03 04:25:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

