

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040220\
 Data File : VX015477.D
 Acq On : 02 Apr 2020 13:28
 Operator : JC/SP
 Sample : L2099-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 239-GW-02-12

Quant Time: Apr 03 08:35:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1158827	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1872953	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1790128	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	939455	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	794674	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
35) Dibromofluoromethane	5.47	113	595466	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
50) Toluene-d8	8.70	98	2261947	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
62) 4-Bromofluorobenzene	11.12	95	960965	53.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.28%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.31	50	6510	0.501	ug/l #	87
16) Acetone	2.42	43	37222	5.301	ug/l	94
51) 4-Methyl-2-Pentanone	8.63	43	11013	0.512	ug/l #	75
52) Toluene	8.77	92	83362	2.675	ug/l	95
68) m/p-Xylenes	10.34	106	5280	0.224	ug/l	80
70) Styrene	10.70	104	9273	0.226	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040220\
Data File : VX015477.D
Acq On : 02 Apr 2020 13:28
Operator : JC/SP
Sample : L2099-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
239-GW-02-12

Quant Time: Apr 03 08:35:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

