

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040220\
 Data File : VX015485.D
 Acq On : 02 Apr 2020 16:29
 Operator : JC/SP
 Sample : PB127948ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB127948ZHE#07

Quant Time: Apr 03 09:03:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1068031	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1733162	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1648018	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	872462	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	741001	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
35) Dibromofluoromethane	5.47	113	554367	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
50) Toluene-d8	8.70	98	2096579	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
62) 4-Bromofluorobenzene	11.12	95	919375	55.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.92%	
Target Compounds						
16) Acetone	2.42	43	49033	7.577	ug/l	98
20) Methylene Chloride	2.84	84	15557	1.287	ug/l	92
43) Isopropyl Acetate	6.42	43	380305	10.835	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040220\
Data File : VX015485.D
Acq On : 02 Apr 2020 16:29
Operator : JC/SP
Sample : PB127948ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB127948ZHE#07

Quant Time: Apr 03 09:03:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

