

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040220\
 Data File : VX015487.D
 Acq On : 02 Apr 2020 17:14
 Operator : JC/SP
 Sample : PB127948ZHE#09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB127948ZHE#09

Quant Time: Apr 03 09:10:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1118099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1818656	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1743007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	940279	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	769546	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
35) Dibromofluoromethane	5.47	113	578027	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
50) Toluene-d8	8.70	98	2208158	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
62) 4-Bromofluorobenzene	11.12	95	947964	54.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.00%	
Target Compounds						
16) Acetone	2.42	43	47337	6.987	ug/l	99
20) Methylene Chloride	2.84	84	14548	1.150	ug/l #	82
43) Isopropyl Acetate	6.42	43	388539	10.549	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040220\
Data File : VX015487.D
Acq On : 02 Apr 2020 17:14
Operator : JC/SP
Sample : PB127948ZHE#09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB127948ZHE#09

Quant Time: Apr 03 09:10:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

